

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052824\
 Data File : PP065250.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 May 2024 10:55
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 29 01:08:35 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP052124.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 21 18:19:41 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.468	3.751	80679888	70834680	51.118	46.660
2) SA Decachlor...	10.206	8.821	73539861	64930129	56.556	49.261
Target Compounds						
21) L5 AR-1248-1	5.632	4.850	16779810	14946896	505.562	495.915
22) L5 AR-1248-2	5.905	5.089	24817845	21166186	487.656	465.614
23) L5 AR-1248-3	6.109	5.131	26656365	22398201	488.496	470.948
24) L5 AR-1248-4	6.512	5.305	28823378	26101954	526.378	472.716
25) L5 AR-1248-5	6.551	5.701	28378113	24820259	527.090	516.888



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052824\
Data File : PP065250.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 May 2024 10:55
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 29 01:08:35 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP052124.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 21 18:19:41 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.468 min
Delta R.T.: 0.000 min
Response: 80679888
Conc: 51.12 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.751 min
Delta R.T.: -0.004 min
Response: 70834680
Conc: 46.66 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.206 min
Delta R.T.: -0.001 min
Response: 73539861
Conc: 56.56 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.821 min
Delta R.T.: -0.015 min
Response: 64930129
Conc: 49.26 ng/ml



#21 AR-1248-1


R.T.: 5.632 min
Delta R.T.: -0.002 min
Response: 16779810
Conc: 505.56 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.850 min
Delta R.T.: -0.007 min
Response: 14946896
Conc: 495.92 ng/ml




#22 AR-1248-2

R.T.: 5.905 min
Delta R.T.: -0.001 min
Response: 24817845
Conc: 487.66 ng/ml



#22 AR-1248-2


R.T.: 5.089 min
Delta R.T.: -0.008 min
Response: 21166186
Conc: 465.61 ng/ml



#23 AR-1248-3


R.T.: 6.109 min
Delta R.T.: -0.002 min
Response: 26656365
Conc: 488.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.131 min
Delta R.T.: -0.008 min
Response: 22398201
Conc: 470.95 ng/ml




#24 AR-1248-4

R.T.: 6.512 min
Delta R.T.: -0.002 min
Response: 28823378
Conc: 526.38 ng/ml



#24 AR-1248-4


R.T.: 5.305 min
Delta R.T.: -0.008 min
Response: 26101954
Conc: 472.72 ng/ml



#25 AR-1248-5

R.T.: 6.551 min
Delta R.T.: -0.002 min
Response: 28378113
Conc: 527.09 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.701 min
Delta R.T.: -0.008 min
Response: 24820259
Conc: 516.89 ng/ml