

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052924\
 Data File : PP065299.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 May 2024 12:19
 Operator : YP\AJ
 Sample : P2622-19DL 20X
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 S02-04ADL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 30 01:27:36 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP052124.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 21 18:19:41 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.468	3.750	1632379	1755636	1.034	1.156
2) SA Decachlor...	10.206	8.816	3215532	2701404	2.473	2.049
Target Compounds						
26) L6 AR-1254-1	6.488	5.659	14288650	15451658	231.428	193.683
27) L6 AR-1254-2	6.703	5.806	22491476	15292581	245.160	216.856
28) L6 AR-1254-3	7.074	6.213	52579641	52373052	565.239	475.205
29) L6 AR-1254-4	7.355	6.440	27821134	22918010	458.933	405.957
30) L6 AR-1254-5	7.792f	6.860	122.5E6	135.6E6	1743.849	1455.693



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052924\
Data File : PP065299.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 May 2024 12:19
Operator : YP\AJ
Sample : P2622-19DL 20X
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
S02-04ADL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 01:27:36 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP052124.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 21 18:19:41 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.468 min
Delta R.T.: 0.000 min
Response: 1632379
Conc: 1.03 ng/ml

Instrument :
ECD_P
ClientSampleId :
S02-04ADL




#1 Tetrachloro-m-xylene

R.T.: 3.750 min
Delta R.T.: -0.005 min
Response: 1755636
Conc: 1.16 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.206 min
Delta R.T.: -0.002 min
Response: 3215532
Conc: 2.47 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.816 min
Delta R.T.: -0.020 min
Response: 2701404
Conc: 2.05 ng/ml



#26 AR-1254-1


R.T.: 6.488 min
Delta R.T.: 0.000 min
Response: 14288650
Conc: 231.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
S02-04ADL




#26 AR-1254-1

R.T.: 5.659 min
Delta R.T.: -0.009 min
Response: 15451658
Conc: 193.68 ng/ml




#27 AR-1254-2

R.T.: 6.703 min
Delta R.T.: -0.003 min
Response: 22491476
Conc: 245.16 ng/ml



#27 AR-1254-2


R.T.: 5.806 min
Delta R.T.: -0.010 min
Response: 15292581
Conc: 216.86 ng/ml



#28 AR-1254-3


R.T.: 7.074 min
Delta R.T.: 0.002 min
Response: 52579641
Conc: 565.24 ng/ml

Instrument :
ECD_P
ClientSampleId :
S02-04ADL




#28 AR-1254-3

R.T.: 6.213 min
Delta R.T.: -0.010 min
Response: 52373052
Conc: 475.20 ng/ml




#29 AR-1254-4

R.T.: 7.355 min
Delta R.T.: -0.002 min
Response: 27821134
Conc: 458.93 ng/ml



#29 AR-1254-4


R.T.: 6.440 min
Delta R.T.: -0.011 min
Response: 22918010
Conc: 405.96 ng/ml



#30 AR-1254-5

R.T.: 7.792 min
Delta R.T.: 0.017 min
Response: 122504350
Conc: 1743.85 ng/ml

Instrument :
ECD_P
ClientSampleId :
S02-04ADL



#30 AR-1254-5

R.T.: 6.860 min
Delta R.T.: -0.013 min
Response: 135641290
Conc: 1455.69 ng/ml