

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052925\
 Data File : PP072454.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 May 2025 20:38
 Operator : YP\AJ
 Sample : Q2154-06 10X
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 BC278406-1-2

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 30 04:26:01 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 24 03:32:20 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.499	3.795	2612887	2257770	1.299	1.413
2) SA Decachlor...	10.204	8.814	2697901	1741971	1.656	1.646
Target Compounds						
26) L6 AR-1254-1	6.500	5.679	23206700	35609617	269.835	356.259 #
27) L6 AR-1254-2	6.716	5.828	36714728	29737106	280.957	344.551
28) L6 AR-1254-3	7.079	6.230	36173089	41941880	274.032	326.090
29) L6 AR-1254-4	7.361	6.458	34200796	29232988	262.054	342.722 #
30) L6 AR-1254-5	7.777	6.876	24893603	31672290	223.769	279.416



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052925\
Data File : PP072454.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 May 2025 20:38
Operator : YP\AJ
Sample : Q2154-06 10X
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
BC278406-1-2

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 04:26:01 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.499 min
Delta R.T.: 0.001 min
Response: 2612887
Conc: 1.30 ng/ml

Instrument :
ECD_P
ClientSampleId :
BC278406-1-2




#1 Tetrachloro-m-xylene

R.T.: 3.795 min
Delta R.T.: 0.001 min
Response: 2257770
Conc: 1.41 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.204 min
Delta R.T.: -0.008 min
Response: 2697901
Conc: 1.66 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.814 min
Delta R.T.: -0.007 min
Response: 1741971
Conc: 1.65 ng/ml



#26 AR-1254-1


R.T.: 6.500 min
Delta R.T.: -0.007 min
Response: 23206700
Conc: 269.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
BC278406-1-2




#26 AR-1254-1

R.T.: 5.679 min
Delta R.T.: -0.006 min
Response: 35609617
Conc: 356.26 ng/ml




#27 AR-1254-2

R.T.: 6.716 min
Delta R.T.: -0.008 min
Response: 36714728
Conc: 280.96 ng/ml



#27 AR-1254-2


R.T.: 5.828 min
Delta R.T.: -0.006 min
Response: 29737106
Conc: 344.55 ng/ml



#28 AR-1254-3


R.T.: 7.079 min
Delta R.T.: -0.009 min
Response: 36173089
Conc: 274.03 ng/ml

Instrument :
ECD_P
ClientSampleId :
BC278406-1-2




#28 AR-1254-3

R.T.: 6.230 min
Delta R.T.: -0.006 min
Response: 41941880
Conc: 326.09 ng/ml




#29 AR-1254-4

R.T.: 7.361 min
Delta R.T.: -0.009 min
Response: 34200796
Conc: 262.05 ng/ml



#29 AR-1254-4


R.T.: 6.458 min
Delta R.T.: -0.007 min
Response: 29232988
Conc: 342.72 ng/ml



#30 AR-1254-5

R.T.: 7.777 min
Delta R.T.: -0.010 min
Response: 24893603
Conc: 223.77 ng/ml

Instrument :
ECD_P
ClientSampleId :
BC278406-1-2



#30 AR-1254-5

R.T.: 6.876 min
Delta R.T.: -0.007 min
Response: 31672290
Conc: 279.42 ng/ml