

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053023\
 Data File : PP058026.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 May 2023 15:35
 Operator : YP\AJ
 Sample : 02968-01
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 34286

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/31/2023
 Supervised By :Ankita Jodhani 05/31/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 30 21:04:23 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 08 19:18:05 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.383	3.624	39784289	35106091	18.135	15.477
2) SA Decachlor...	10.225	8.698	20645418	20999988	18.218	13.378 #
Target Compounds						
31) L7 AR-1260-1	7.189	6.228	12122405	15381701	153.649	139.905
32) L7 AR-1260-2	7.450	6.421	22777364	22512508	284.297	177.132 #
33) L7 AR-1260-3	7.813	6.572	15129774	13715192	265.554m	111.850 #
34) L7 AR-1260-4	8.040	7.049	21739698	14192649	327.192	162.423 #
35) L7 AR-1260-5	8.366	7.295	35452027	36363507	330.158	196.876 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053023\
Data File : PP058026.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2023 15:35
Operator : YP\AJ
Sample : 02968-01
Misc :
ALS Vial : 16 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
34286


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/31/2023
Supervised By :Ankita Jodhani 05/31/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 21:04:23 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050823.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 08 19:18:05 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.383 min
Delta R.T.: -0.007 min
Response: 39784289
Conc: 18.13 ng/ml

Instrument :
ECD_P
ClientSampleId :
34286


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/31/2023
Supervised By :Ankita Jodhani 05/31/2023




#1 Tetrachloro-m-xylene

R.T.: 3.624 min
Delta R.T.: 0.000 min
Response: 35106091
Conc: 15.48 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.225 min
Delta R.T.: 0.002 min
Response: 20645418
Conc: 18.22 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.698 min
Delta R.T.: 0.013 min
Response: 20999988
Conc: 13.38 ng/ml




#31 AR-1260-1

R.T.: 7.189 min
Delta R.T.: -0.002 min
Response: 12122405
Conc: 153.65 ng/ml

Instrument :
ECD_P
ClientSampleId :
34286


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/31/2023
Supervised By :Ankita Jodhani 05/31/2023




#31 AR-1260-1

R.T.: 6.228 min
Delta R.T.: 0.009 min
Response: 15381701
Conc: 139.90 ng/ml




#32 AR-1260-2

R.T.: 7.450 min
Delta R.T.: 0.000 min
Response: 22777364
Conc: 284.30 ng/ml



#32 AR-1260-2

R.T.: 6.421 min
Delta R.T.: 0.012 min
Response: 22512508
Conc: 177.13 ng/ml




#33 AR-1260-3

R.T.: 7.813 min
Delta R.T.: 0.000 min
Response: 15129774
Conc: 265.55 ng/ml

Instrument :
ECD_P
ClientSampleId :
34286


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/31/2023
Supervised By :Ankita Jodhani 05/31/2023




#33 AR-1260-3

R.T.: 6.572 min
Delta R.T.: 0.008 min
Response: 13715192
Conc: 111.85 ng/ml




#34 AR-1260-4

R.T.: 8.040 min
Delta R.T.: 0.000 min
Response: 21739698
Conc: 327.19 ng/ml



#34 AR-1260-4

R.T.: 7.049 min
Delta R.T.: 0.010 min
Response: 14192649
Conc: 162.42 ng/ml




#35 AR-1260-5

R.T.: 8.366 min
Delta R.T.: 0.002 min
Response: 35452027
Conc: 330.16 ng/ml

Instrument :
ECD_P
ClientSampleId :
34286

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/31/2023
Supervised By :Ankita Jodhani 05/31/2023



#35 AR-1260-5

R.T.: 7.295 min
Delta R.T.: 0.013 min
Response: 36363507
Conc: 196.88 ng/ml