

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053023\
 Data File : PP058052.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 May 2023 04:40
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 31 05:00:57 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 08 19:18:05 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.382	3.622	97227241	108.9E6	44.318	48.019
2) SA Decachlor...	10.205	8.677	60337523	67882875	53.243	43.245
Target Compounds						
21) L5 AR-1248-1	5.561	4.716	18244324	21642801	433.533	420.585
22) L5 AR-1248-2	5.837	4.955	28431362	30059900	436.061	418.644
23) L5 AR-1248-3	6.043	4.997	31295725	31816880	450.676	420.531
24) L5 AR-1248-4	6.450	5.170	29641561	37481884	463.796	420.193
25) L5 AR-1248-5	6.489	5.567	30858329	34551649	480.966	415.574



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053023\
Data File : PP058052.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 May 2023 04:40
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 31 05:00:57 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050823.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 08 19:18:05 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.382 min
Delta R.T.: -0.008 min
Response: 97227241
Conc: 44.32 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.622 min
Delta R.T.: -0.002 min
Response: 108924080
Conc: 48.02 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.205 min
Delta R.T.: -0.018 min
Response: 60337523
Conc: 53.24 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.677 min
Delta R.T.: -0.009 min
Response: 67882875
Conc: 43.24 ng/ml



#21 AR-1248-1


R.T.: 5.561 min
Delta R.T.: -0.008 min
Response: 18244324
Conc: 433.53 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.716 min
Delta R.T.: -0.004 min
Response: 21642801
Conc: 420.58 ng/ml




#22 AR-1248-2

R.T.: 5.837 min
Delta R.T.: -0.009 min
Response: 28431362
Conc: 436.06 ng/ml



#22 AR-1248-2


R.T.: 4.955 min
Delta R.T.: -0.005 min
Response: 30059900
Conc: 418.64 ng/ml



#23 AR-1248-3


R.T.: 6.043 min
Delta R.T.: -0.009 min
Response: 31295725
Conc: 450.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.997 min
Delta R.T.: -0.005 min
Response: 31816880
Conc: 420.53 ng/ml




#24 AR-1248-4

R.T.: 6.450 min
Delta R.T.: -0.009 min
Response: 29641561
Conc: 463.80 ng/ml



#24 AR-1248-4


R.T.: 5.170 min
Delta R.T.: -0.005 min
Response: 37481884
Conc: 420.19 ng/ml



#25 AR-1248-5

R.T.: 6.489 min
Delta R.T.: -0.009 min
Response: 30858329
Conc: 480.97 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.567 min
Delta R.T.: -0.005 min
Response: 34551649
Conc: 415.57 ng/ml