

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
 Data File : PP072463.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 May 2025 10:57
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 30 11:14:25 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 24 03:32:20 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.502	3.797	98213000	83714987	48.843	52.376
2) SA Decachlor...	10.210	8.818	87502701	60026703	53.711	56.704
Target Compounds						
21) L5 AR-1248-1	5.654	4.879	21947537	20864221	453.504	481.612
22) L5 AR-1248-2	5.926	5.115	29722513	27487945	490.146	486.708
23) L5 AR-1248-3	6.128	5.158	33211488	28668110	501.835	485.020
24) L5 AR-1248-4	6.527	5.330	41928908	33698686	479.779	483.809
25) L5 AR-1248-5	6.566	5.722	40881993	33019689	492.402	473.472



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
Data File : PP072463.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2025 10:57
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 11:14:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.502 min
Delta R.T.: 0.004 min
Response: 98213000
Conc: 48.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.797 min
Delta R.T.: 0.003 min
Response: 83714987
Conc: 52.38 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.210 min
Delta R.T.: -0.002 min
Response: 87502701
Conc: 53.71 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.818 min
Delta R.T.: -0.003 min
Response: 60026703
Conc: 56.70 ng/ml



#21 AR-1248-1


R.T.: 5.654 min
Delta R.T.: -0.004 min
Response: 21947537
Conc: 453.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.879 min
Delta R.T.: -0.002 min
Response: 20864221
Conc: 481.61 ng/ml




#22 AR-1248-2

R.T.: 5.926 min
Delta R.T.: -0.005 min
Response: 29722513
Conc: 490.15 ng/ml



#22 AR-1248-2


R.T.: 5.115 min
Delta R.T.: -0.003 min
Response: 27487945
Conc: 486.71 ng/ml



#23 AR-1248-3


R.T.: 6.128 min
Delta R.T.: -0.005 min
Response: 33211488
Conc: 501.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.158 min
Delta R.T.: -0.003 min
Response: 28668110
Conc: 485.02 ng/ml




#24 AR-1248-4

R.T.: 6.527 min
Delta R.T.: -0.006 min
Response: 41928908
Conc: 479.78 ng/ml



#24 AR-1248-4


R.T.: 5.330 min
Delta R.T.: -0.003 min
Response: 33698686
Conc: 483.81 ng/ml



#25 AR-1248-5

R.T.: 6.566 min
Delta R.T.: -0.005 min
Response: 40881993
Conc: 492.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.722 min
Delta R.T.: -0.003 min
Response: 33019689
Conc: 473.47 ng/ml