

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
 Data File : PP072479.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 May 2025 17:04
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 30 19:10:55 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 24 03:32:20 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.503	3.796	98838806	86661465	49.154	54.219
2)	SA Decachlor...	10.211	8.817	85027080	59122875	52.192	55.850
Target Compounds							
26)	L6 AR-1254-1	6.505	5.681	39144409	50070268	455.150	500.932
27)	L6 AR-1254-2	6.721	5.829	59294459	42636080	453.747	494.005
28)	L6 AR-1254-3	7.085	6.231	62344103	65912235	472.293	512.454
29)	L6 AR-1254-4	7.367	6.460	57990739	42514734	444.337	498.435
30)	L6 AR-1254-5	7.784	6.877	54166534	56946859	486.905	502.391



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
Data File : PP072479.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2025 17:04
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 19:10:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.503 min
Delta R.T.: 0.005 min
Response: 98838806
Conc: 49.15 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.796 min
Delta R.T.: 0.003 min
Response: 86661465
Conc: 54.22 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.211 min
Delta R.T.: 0.000 min
Response: 85027080
Conc: 52.19 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.817 min
Delta R.T.: -0.004 min
Response: 59122875
Conc: 55.85 ng/ml



#26 AR-1254-1


R.T.: 6.505 min
Delta R.T.: -0.002 min
Response: 39144409
Conc: 455.15 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.681 min
Delta R.T.: -0.005 min
Response: 50070268
Conc: 500.93 ng/ml




#27 AR-1254-2

R.T.: 6.721 min
Delta R.T.: -0.003 min
Response: 59294459
Conc: 453.75 ng/ml



#27 AR-1254-2


R.T.: 5.829 min
Delta R.T.: -0.005 min
Response: 42636080
Conc: 494.01 ng/ml



#28 AR-1254-3


R.T.: 7.085 min
Delta R.T.: -0.003 min
Response: 62344103
Conc: 472.29 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.231 min
Delta R.T.: -0.005 min
Response: 65912235
Conc: 512.45 ng/ml




#29 AR-1254-4

R.T.: 7.367 min
Delta R.T.: -0.003 min
Response: 57990739
Conc: 444.34 ng/ml



#29 AR-1254-4


R.T.: 6.460 min
Delta R.T.: -0.005 min
Response: 42514734
Conc: 498.44 ng/ml



#30 AR-1254-5

R.T.: 7.784 min
Delta R.T.: -0.003 min
Response: 54166534
Conc: 486.90 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.877 min
Delta R.T.: -0.006 min
Response: 56946859
Conc: 502.39 ng/ml