

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
 Data File : PP072492.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 May 2025 21:42
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 30 22:57:56 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 24 03:32:20 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.503	3.796	94068035	78504123	46.781	49.115
2)	SA Decachlor...	10.211	8.817	75068011	52244088	46.079	49.352
Target Compounds							
16)	L4 AR-1242-1	5.656	4.878	27162300	26338827	433.921	505.462
17)	L4 AR-1242-2	5.678	4.897	41479189	37648682	477.982	512.253
18)	L4 AR-1242-3	5.740	5.073	25188633	20888725	457.427	533.890
19)	L4 AR-1242-4	5.838	5.157	21134619	19951182	476.552	530.859
20)	L4 AR-1242-5	6.568	5.680	21950504	25079838	430.530	519.707



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
Data File : PP072492.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2025 21:42
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 22:57:56 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.503 min
Delta R.T.: 0.005 min
Response: 94068035
Conc: 46.78 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.796 min
Delta R.T.: 0.003 min
Response: 78504123
Conc: 49.12 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.211 min
Delta R.T.: -0.001 min
Response: 75068011
Conc: 46.08 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.817 min
Delta R.T.: -0.004 min
Response: 52244088
Conc: 49.35 ng/ml



#16 AR-1242-1


R.T.: 5.656 min
Delta R.T.: 0.000 min
Response: 27162300
Conc: 433.92 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.878 min
Delta R.T.: -0.002 min
Response: 26338827
Conc: 505.46 ng/ml




#17 AR-1242-2

R.T.: 5.678 min
Delta R.T.: 0.000 min
Response: 41479189
Conc: 477.98 ng/ml



#17 AR-1242-2


R.T.: 4.897 min
Delta R.T.: -0.002 min
Response: 37648682
Conc: 512.25 ng/ml



#18 AR-1242-3


R.T.: 5.740 min
Delta R.T.: 0.000 min
Response: 25188633
Conc: 457.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 5.073 min
Delta R.T.: -0.002 min
Response: 20888725
Conc: 533.89 ng/ml




#19 AR-1242-4

R.T.: 5.838 min
Delta R.T.: 0.000 min
Response: 21134619
Conc: 476.55 ng/ml



#19 AR-1242-4


R.T.: 5.157 min
Delta R.T.: -0.003 min
Response: 19951182
Conc: 530.86 ng/ml



#20 AR-1242-5

R.T.: 6.568 min
Delta R.T.: 0.000 min
Response: 21950504
Conc: 430.53 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.680 min
Delta R.T.: -0.003 min
Response: 25079838
Conc: 519.71 ng/ml