

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053122\
 Data File : PP048217.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 May 2022 10:48
 Operator : YP\AJ
 Sample : N2932-02RE
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 P016-SS002-2430-01RE

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/01/2022
 Supervised By :Ankita Jodhani 06/01/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 01 07:09:40 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP052722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 27 23:53:39 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.036	3.217	48640677	41403304	20.771	21.009
2) SA Decachlor...	10.390	8.652	86833015	65658418	44.676	40.618
Target Compounds						
31) L7 AR-1260-1	7.094	5.981	2467615	1530474	26.438	17.879 #
32) L7 AR-1260-2	7.381	6.188	6409546	3594660	57.338	34.873 #
33) L7 AR-1260-3	7.779	6.357	6801886	2773784	79.409	28.012 #
34) L7 AR-1260-4	8.033	6.870	7924735	5574062	76.378m	66.955
35) L7 AR-1260-5	8.393	7.134	14591164	11587492	70.708	58.401

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053122\
Data File : PP048217.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 May 2022 10:48
Operator : YP\AJ
Sample : N2932-02RE
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
P016-SS002-2430-01RE


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/01/2022
Supervised By :Ankita Jodhani 06/01/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 01 07:09:40 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP052722.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 27 23:53:39 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.036 min
Delta R.T.: 0.000 min
Response: 48640677
Conc: 20.77 ng/ml

Instrument :
ECD_P
ClientSampleId :
P016-SS002-2430-01RE


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/01/2022
Supervised By :Ankita Jodhani 06/01/2022




#1 Tetrachloro-m-xylene

R.T.: 3.217 min
Delta R.T.: -0.006 min
Response: 41403304
Conc: 21.01 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.390 min
Delta R.T.: -0.002 min
Response: 86833015
Conc: 44.68 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.652 min
Delta R.T.: -0.017 min
Response: 65658418
Conc: 40.62 ng/ml




#33 AR-1260-3

R.T.: 7.779 min
Delta R.T.: -0.002 min
Response: 6801886
Conc: 79.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
P016-SS002-2430-01RE


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/01/2022
Supervised By :Ankita Jodhani 06/01/2022




#33 AR-1260-3

R.T.: 6.357 min
Delta R.T.: -0.008 min
Response: 2773784
Conc: 28.01 ng/ml




#34 AR-1260-4

R.T.: 8.033 min
Delta R.T.: 0.000 min
Response: 7924735
Conc: 76.38 ng/ml m



#34 AR-1260-4

R.T.: 6.870 min
Delta R.T.: -0.014 min
Response: 5574062
Conc: 66.96 ng/ml




#35 AR-1260-5

R.T.: 8.393 min
Delta R.T.: -0.002 min
Response: 14591164
Conc: 70.71 ng/ml

Instrument :
ECD_P
ClientSampleId :
P016-SS002-2430-01RE

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/01/2022
Supervised By :Ankita Jodhani 06/01/2022



#35 AR-1260-5

R.T.: 7.134 min
Delta R.T.: -0.014 min
Response: 11587492
Conc: 58.40 ng/ml