

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053122\
 Data File : PP048227.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 May 2022 13:21
 Operator : YP\AJ
 Sample : N2932-16
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 P016-SS005-0006-01

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/01/2022
 Supervised By :Ankita Jodhani 06/01/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 01 07:14:57 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP052722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 27 23:53:39 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.036	3.238	62475449	50867571	26.679	25.811
2) SA Decachlor...	10.387	8.673	56284044	45893029	28.958	28.391
Target Compounds						
31) L7 AR-1260-1	7.095	6.003	942344	591780	10.096m	6.913 #
32) L7 AR-1260-2	7.382	6.209	3143064	3061591	28.117	29.702
33) L7 AR-1260-3	7.779	6.382f	2616551	1319447	30.547	13.325 #
34) L7 AR-1260-4	8.033	6.891	1689100	1835453	16.279m	22.047 #
35) L7 AR-1260-5	8.396	7.156	4733129	6378350	22.937m	32.147 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053122\
Data File : PP048227.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 May 2022 13:21
Operator : YP\AJ
Sample : N2932-16
Misc :
ALS Vial : 17 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
P016-SS005-0006-01


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/01/2022
Supervised By :Ankita Jodhani 06/01/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 01 07:14:57 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP052722.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 27 23:53:39 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.036 min
Delta R.T.: -0.001 min
Response: 62475449
Conc: 26.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
P016-SS005-0006-01


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/01/2022
Supervised By :Ankita Jodhani 06/01/2022




#1 Tetrachloro-m-xylene

R.T.: 3.238 min
Delta R.T.: 0.015 min
Response: 50867571
Conc: 25.81 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.387 min
Delta R.T.: -0.005 min
Response: 56284044
Conc: 28.96 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.673 min
Delta R.T.: 0.004 min
Response: 45893029
Conc: 28.39 ng/ml




#31 AR-1260-1

R.T.: 7.095 min
Delta R.T.: 0.000 min
Response: 942344
Conc: 10.10 ng/ml

Instrument :
ECD_P
ClientSampleId :
P016-SS005-0006-01


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/01/2022
Supervised By :Ankita Jodhani 06/01/2022




#31 AR-1260-1

R.T.: 6.003 min
Delta R.T.: 0.010 min
Response: 591780
Conc: 6.91 ng/ml




#32 AR-1260-2

R.T.: 7.382 min
Delta R.T.: -0.002 min
Response: 3143064
Conc: 28.12 ng/ml



#32 AR-1260-2

R.T.: 6.209 min
Delta R.T.: 0.010 min
Response: 3061591
Conc: 29.70 ng/ml




#33 AR-1260-3

R.T.: 7.779 min
Delta R.T.: -0.002 min
Response: 2616551
Conc: 30.55 ng/ml

Instrument :
ECD_P
ClientSampleId :
P016-SS005-0006-01


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/01/2022
Supervised By :Ankita Jodhani 06/01/2022




#33 AR-1260-3

R.T.: 6.382 min
Delta R.T.: 0.017 min
Response: 1319447
Conc: 13.32 ng/ml




#34 AR-1260-4

R.T.: 8.033 min
Delta R.T.: 0.000 min
Response: 1689100
Conc: 16.28 ng/ml m



#34 AR-1260-4

R.T.: 6.891 min
Delta R.T.: 0.007 min
Response: 1835453
Conc: 22.05 ng/ml




#35 AR-1260-5

R.T.: 8.396 min
Delta R.T.: 0.000 min
Response: 4733129
Conc: 22.94 ng/ml

Instrument :
ECD_P
ClientSampleId :
P016-SS005-0006-01

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/01/2022
Supervised By :Ankita Jodhani 06/01/2022



#35 AR-1260-5

R.T.: 7.156 min
Delta R.T.: 0.008 min
Response: 6378350
Conc: 32.15 ng/ml