

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053122\
 Data File : PP048252.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 May 2022 20:10
 Operator : YP\AJ
 Sample : N2954-08
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 P044-SS004-1824-01

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 01 11:03:02 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP052722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 27 23:53:39 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.035	3.216	55970364	45716576	23.901	23.197
2)	SA Decachlor...	10.386	8.648	38085446	31177660	19.595	19.287
Target Compounds							
31)	L7 AR-1260-1	7.093	5.979	222.6E6	208.1E6	2385.143	2431.534
32)	L7 AR-1260-2	7.381	6.185	302.0E6	296.9E6	2701.914	2880.804
33)	L7 AR-1260-3	7.777	6.352	192.7E6	117.7E6	2250.053	1188.937 #
34)	L7 AR-1260-4	8.030	6.867	145.9E6	114.3E6	1406.481	1373.200
35)	L7 AR-1260-5	8.392	7.133	273.7E6	284.9E6	1326.444	1435.694



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053122\
Data File : PP048252.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 May 2022 20:10
Operator : YP\AJ
Sample : N2954-08
Misc :
ALS Vial : 41 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
P044-SS004-1824-01

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 01 11:03:02 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP052722.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 27 23:53:39 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.035 min
Delta R.T.: -0.002 min
Response: 55970364
Conc: 23.90 ng/ml

Instrument :
ECD_P
ClientSampleId :
P044-SS004-1824-01




#1 Tetrachloro-m-xylene

R.T.: 3.216 min
Delta R.T.: -0.007 min
Response: 45716576
Conc: 23.20 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.386 min
Delta R.T.: -0.007 min
Response: 38085446
Conc: 19.60 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.648 min
Delta R.T.: -0.012 min
Response: 31177660
Conc: 19.29 ng/ml



#31 AR-1260-1


R.T.: 7.093 min
Delta R.T.: -0.003 min
Response: 222620497
Conc: 2385.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
P044-SS004-1824-01




#31 AR-1260-1

R.T.: 5.979 min
Delta R.T.: -0.014 min
Response: 208147743
Conc: 2431.53 ng/ml




#32 AR-1260-2

R.T.: 7.381 min
Delta R.T.: -0.004 min
Response: 302035497
Conc: 2701.91 ng/ml



#32 AR-1260-2


R.T.: 6.185 min
Delta R.T.: -0.014 min
Response: 296946047
Conc: 2880.80 ng/ml



#33 AR-1260-3


R.T.: 7.777 min
Delta R.T.: -0.004 min
Response: 192732289
Conc: 2250.05 ng/ml

Instrument :
ECD_P
ClientSampleId :
P044-SS004-1824-01




#33 AR-1260-3

R.T.: 6.352 min
Delta R.T.: -0.013 min
Response: 117731463
Conc: 1188.94 ng/ml




#34 AR-1260-4

R.T.: 8.030 min
Delta R.T.: -0.003 min
Response: 145932098
Conc: 1406.48 ng/ml



#34 AR-1260-4


R.T.: 6.867 min
Delta R.T.: -0.008 min
Response: 114319788
Conc: 1373.20 ng/ml



#35 AR-1260-5

R.T.: 8.392 min
Delta R.T.: -0.004 min
Response: 273721065
Conc: 1326.44 ng/ml

Instrument :
ECD_P
ClientSampleId :
P044-SS004-1824-01



#35 AR-1260-5

R.T.: 7.133 min
Delta R.T.: -0.007 min
Response: 284859044
Conc: 1435.69 ng/ml