

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053122\
 Data File : PP048267.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Jun 2022 05:31
 Operator : YP\AJ
 Sample : N2952-09
 Misc :
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 P037-SS001-1218-01

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/01/2022
 Supervised By :Ankita Jodhani 06/01/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 01 11:44:52 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP052722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 27 23:53:39 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.036	3.216	55866062	45451709	23.856	23.063
2) SA Decachlor...	10.387	8.648	41663348	33704975	21.436	20.851
Target Compounds						
31) L7 AR-1260-1	7.095	5.980	2129943	835975	22.820m	9.766 #
32) L7 AR-1260-2	7.381	6.185	2172147	1704901	19.431m	16.540
33) L7 AR-1260-3	7.779	6.353	1968015	817080	22.976m	8.251 #
34) L7 AR-1260-4	8.029	6.869	2004807	1365867	19.322m	16.407
35) L7 AR-1260-5	8.393	7.132	5331201	4083539	25.835	20.581

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053122\
Data File : PP048267.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Jun 2022 05:31
Operator : YP\AJ
Sample : N2952-09
Misc :
ALS Vial : 54 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
P037-SS001-1218-01


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/01/2022
Supervised By :Ankita Jodhani 06/01/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 01 11:44:52 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP052722.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 27 23:53:39 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.036 min
Delta R.T.: -0.001 min
Response: 55866062
Conc: 23.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
P037-SS001-1218-01


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/01/2022
Supervised By :Ankita Jodhani 06/01/2022




#1 Tetrachloro-m-xylene

R.T.: 3.216 min
Delta R.T.: -0.007 min
Response: 45451709
Conc: 23.06 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.387 min
Delta R.T.: -0.005 min
Response: 41663348
Conc: 21.44 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.648 min
Delta R.T.: -0.012 min
Response: 33704975
Conc: 20.85 ng/ml




#31 AR-1260-1

R.T.: 7.095 min
Delta R.T.: -0.002 min
Response: 2129943
Conc: 22.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
P037-SS001-1218-01


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/01/2022
Supervised By :Ankita Jodhani 06/01/2022




#31 AR-1260-1

R.T.: 5.980 min
Delta R.T.: -0.005 min
Response: 835975
Conc: 9.77 ng/ml




#32 AR-1260-2

R.T.: 7.381 min
Delta R.T.: -0.003 min
Response: 2172147
Conc: 19.43 ng/ml m



#32 AR-1260-2

R.T.: 6.185 min
Delta R.T.: -0.005 min
Response: 1704901
Conc: 16.54 ng/ml




#33 AR-1260-3

R.T.: 7.779 min
Delta R.T.: -0.002 min
Response: 1968015
Conc: 22.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
P037-SS001-1218-01


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/01/2022
Supervised By :Ankita Jodhani 06/01/2022




#33 AR-1260-3

R.T.: 6.353 min
Delta R.T.: -0.002 min
Response: 817080
Conc: 8.25 ng/ml




#34 AR-1260-4

R.T.: 8.029 min
Delta R.T.: -0.005 min
Response: 2004807
Conc: 19.32 ng/ml m



#34 AR-1260-4

R.T.: 6.869 min
Delta R.T.: -0.006 min
Response: 1365867
Conc: 16.41 ng/ml




#35 AR-1260-5

R.T.: 8.393 min
Delta R.T.: -0.003 min
Response: 5331201
Conc: 25.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
P037-SS001-1218-01

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/01/2022
Supervised By :Ankita Jodhani 06/01/2022



#35 AR-1260-5

R.T.: 7.132 min
Delta R.T.: -0.008 min
Response: 4083539
Conc: 20.58 ng/ml