

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053123\
 Data File : PP058098.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 May 2023 23:50
 Operator : YP\AJ
 Sample : 02959-05
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 60062

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 01 09:10:23 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 08 19:18:05 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.382	3.624	36830917	45714662	16.788	20.153
2) SA Decachlor...	10.202	8.676	25348859	25659680	22.368	16.346 #
Target Compounds						
16) L4 AR-1242-1	5.562	4.718	311.5E6	354.5E6	6539.271	6089.171
17) L4 AR-1242-2	5.584	4.738	434.4E6	463.1E6	6352.521	5801.480
18) L4 AR-1242-3	5.647	4.916	240.6E6	249.4E6	5524.970	5318.414
19) L4 AR-1242-4	5.746	5.000	215.1E6	344.3E6	6103.225	7421.202
20) L4 AR-1242-5	6.490	5.528	259.2E6	350.4E6	7999.944	6267.251
26) L6 AR-1254-1	6.423	5.528	188.6E6	350.4E6	2810.527	3138.117
27) L6 AR-1254-2	6.648	5.677	273.0E6	124.9E6	2885.885	1251.300 #
28) L6 AR-1254-3	7.014	6.084	205.0E6	246.9E6	2250.567	1578.339 #
29) L6 AR-1254-4	7.303	6.314	221.3E6	248.1E6	3714.700	2689.136 #
30) L6 AR-1254-5	7.725	6.734	147.3E6	157.8E6	2144.929	1089.983 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053123\
Data File : PP058098.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 May 2023 23:50
Operator : YP\AJ
Sample : 02959-05
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
60062

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 01 09:10:23 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050823.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 08 19:18:05 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.382 min
Delta R.T.: -0.008 min
Response: 36830917
Conc: 16.79 ng/ml

Instrument :
ECD_P
ClientSampleId :
60062




#1 Tetrachloro-m-xylene

R.T.: 3.624 min
Delta R.T.: 0.000 min
Response: 45714662
Conc: 20.15 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.202 min
Delta R.T.: -0.021 min
Response: 25348859
Conc: 22.37 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.676 min
Delta R.T.: -0.010 min
Response: 25659680
Conc: 16.35 ng/ml



#16 AR-1242-1


R.T.: 5.562 min
Delta R.T.: -0.006 min
Response: 311517673
Conc: 6539.27 ng/ml

Instrument :
ECD_P
ClientSampleId :
60062




#16 AR-1242-1

R.T.: 4.718 min
Delta R.T.: -0.002 min
Response: 354456852
Conc: 6089.17 ng/ml




#17 AR-1242-2

R.T.: 5.584 min
Delta R.T.: -0.007 min
Response: 434429073
Conc: 6352.52 ng/ml



#17 AR-1242-2


R.T.: 4.738 min
Delta R.T.: -0.002 min
Response: 463063760
Conc: 5801.48 ng/ml



#18 AR-1242-3


R.T.: 5.647 min
Delta R.T.: -0.007 min
Response: 240559579
Conc: 5524.97 ng/ml

Instrument :
ECD_P
ClientSampleId :
60062




#18 AR-1242-3

R.T.: 4.916 min
Delta R.T.: -0.002 min
Response: 249423931
Conc: 5318.41 ng/ml




#19 AR-1242-4

R.T.: 5.746 min
Delta R.T.: -0.007 min
Response: 215059369
Conc: 6103.22 ng/ml



#19 AR-1242-4


R.T.: 5.000 min
Delta R.T.: -0.002 min
Response: 344346545
Conc: 7421.20 ng/ml



#20 AR-1242-5


R.T.: 6.490 min
Delta R.T.: -0.007 min
Response: 259208471
Conc: 7999.94 ng/ml

Instrument :
ECD_P
ClientSampleId :
60062




#20 AR-1242-5

R.T.: 5.528 min
Delta R.T.: -0.002 min
Response: 350400130
Conc: 6267.25 ng/ml




#26 AR-1254-1

R.T.: 6.423 min
Delta R.T.: -0.010 min
Response: 188636584
Conc: 2810.53 ng/ml



#26 AR-1254-1


R.T.: 5.528 min
Delta R.T.: -0.003 min
Response: 350400130
Conc: 3138.12 ng/ml



#27 AR-1254-2


R.T.: 6.648 min
Delta R.T.: -0.006 min
Response: 272999688
Conc: 2885.88 ng/ml

Instrument :
ECD_P
ClientSampleId :
60062




#27 AR-1254-2

R.T.: 5.677 min
Delta R.T.: -0.003 min
Response: 124884232
Conc: 1251.30 ng/ml




#28 AR-1254-3

R.T.: 7.014 min
Delta R.T.: -0.009 min
Response: 205000814
Conc: 2250.57 ng/ml



#28 AR-1254-3


R.T.: 6.084 min
Delta R.T.: -0.004 min
Response: 246935273
Conc: 1578.34 ng/ml



#29 AR-1254-4


R.T.: 7.303 min
Delta R.T.: -0.009 min
Response: 221255352
Conc: 3714.70 ng/ml

Instrument :
ECD_P
ClientSampleId :
60062




#29 AR-1254-4

R.T.: 6.314 min
Delta R.T.: -0.005 min
Response: 248078134
Conc: 2689.14 ng/ml



#30 AR-1254-5

R.T.: 7.725 min
Delta R.T.: -0.011 min
Response: 147253585
Conc: 2144.93 ng/ml



#30 AR-1254-5

R.T.: 6.734 min
Delta R.T.: -0.006 min
Response: 157785754
Conc: 1089.98 ng/ml