

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060122\
 Data File : PP048298.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Jun 2022 15:20
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/02/2022
 Supervised By :Ankita Jodhani 06/02/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 02 04:10:16 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP052722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 27 23:53:39 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.035	3.215	116.6E6	93527766	49.772	47.457
2) SA Decachlor...	10.384	8.645	93590363	78713161	48.153	48.694
Target Compounds						
3) L1 AR-1016-1	5.322	4.375	40528177	31634478	480.598	470.350m
4) L1 AR-1016-2	5.347	4.395	58197101	45394571	491.877	487.429
5) L1 AR-1016-3	5.413	4.582	36864852	24519052	545.533	478.631
6) L1 AR-1016-4	5.522	4.631	30633175	19811552	496.963	519.977
7) L1 AR-1016-5	5.847	4.858	31521224	21451634	559.590	449.291
31) L7 AR-1260-1	7.092	5.977	51551515	46275623	552.320	540.581
32) L7 AR-1260-2	7.379	6.183	58955036	56644195	527.393	549.530
33) L7 AR-1260-3	7.776	6.349	43226980	52290715	504.653	528.069
34) L7 AR-1260-4	8.029	6.866	53094597	40847017	511.721	490.651
35) L7 AR-1260-5	8.391	7.131	94945995	95167729	460.106	479.647

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060122\
Data File : PP048298.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Jun 2022 15:20
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/02/2022
Supervised By :Ankita Jodhani 06/02/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 02 04:10:16 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP052722.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 27 23:53:39 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

