

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060122\
 Data File : PP048282.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Jun 2022 10:07
 Operator : YP\AJ
 Sample : N3109-03
 Misc :
 ALS Vial : 65 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 BB-9

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/02/2022
 Supervised By :Ankita Jodhani 06/02/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 02 10:24:33 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP052722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 27 23:53:39 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.036	3.216	51013038	41562087	21.784	21.089
2) SA Decachlor...	10.386	8.648	30561507	24889453	15.724	15.397
Target Compounds						
26) L6 AR-1254-1	6.268	5.242	3519874	1685681	48.486	19.229 #
27) L6 AR-1254-2	6.509	5.400	4919084	1483909	41.341	19.750 #
28) L6 AR-1254-3	6.919	5.859	3848904	2146858	32.400	17.497 #
29) L6 AR-1254-4	7.231	6.087	1457785	581303	16.162m	7.742 #
30) L6 AR-1254-5	7.699	6.543	3575647	15854408	38.587m	156.607 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060122\
Data File : PP048282.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Jun 2022 10:07
Operator : YP\AJ
Sample : N3109-03
Misc :
ALS Vial : 65 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
BB-9


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/02/2022
Supervised By :Ankita Jodhani 06/02/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 02 10:24:33 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP052722.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 27 23:53:39 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.036 min
Delta R.T.: -0.001 min
Response: 51013038
Conc: 21.78 ng/ml

Instrument :
ECD_P
ClientSampleId :
BB-9


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/02/2022
Supervised By :Ankita Jodhani 06/02/2022




#1 Tetrachloro-m-xylene

R.T.: 3.216 min
Delta R.T.: -0.007 min
Response: 41562087
Conc: 21.09 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.386 min
Delta R.T.: -0.006 min
Response: 30561507
Conc: 15.72 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.648 min
Delta R.T.: -0.012 min
Response: 24889453
Conc: 15.40 ng/ml




#26 AR-1254-1

R.T.: 6.268 min
Delta R.T.: 0.000 min
Response: 3519874
Conc: 48.49 ng/ml

Instrument :
ECD_P
ClientSampleId :
BB-9


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/02/2022
Supervised By :Ankita Jodhani 06/02/2022




#26 AR-1254-1

R.T.: 5.242 min
Delta R.T.: -0.008 min
Response: 1685681
Conc: 19.23 ng/ml




#27 AR-1254-2

R.T.: 6.509 min
Delta R.T.: -0.001 min
Response: 4919084
Conc: 41.34 ng/ml



#27 AR-1254-2

R.T.: 5.400 min
Delta R.T.: -0.011 min
Response: 1483909
Conc: 19.75 ng/ml




#28 AR-1254-3

R.T.: 6.919 min
Delta R.T.: 0.002 min
Response: 3848904
Conc: 32.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
BB-9


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/02/2022
Supervised By :Ankita Jodhani 06/02/2022




#28 AR-1254-3

R.T.: 5.859 min
Delta R.T.: 0.011 min
Response: 2146858
Conc: 17.50 ng/ml




#29 AR-1254-4

R.T.: 7.231 min
Delta R.T.: -0.006 min
Response: 1457785
Conc: 16.16 ng/ml m



#29 AR-1254-4

R.T.: 6.087 min
Delta R.T.: -0.011 min
Response: 581303
Conc: 7.74 ng/ml




#30 AR-1254-5

R.T.: 7.699 min
Delta R.T.: -0.004 min
Response: 3575647
Conc: 38.59 ng/ml

Instrument :
ECD_P
ClientSampleId :
BB-9

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/02/2022
Supervised By :Ankita Jodhani 06/02/2022



#30 AR-1254-5

R.T.: 6.543 min
Delta R.T.: -0.012 min
Response: 15854408
Conc: 156.61 ng/ml