

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060123\
 Data File : PP058171.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Jun 2023 06:59
 Operator : YP\AJ
 Sample : 02999-02
 Misc :
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 50010

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/02/2023
 Supervised By :Ankita Jodhani 06/02/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 02 07:16:29 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 08 19:18:05 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.382	3.621	55312220	56217841	25.213	24.784
2) SA Decachlor...	10.204	8.678	33628120	38356551	29.674	24.435
Target Compounds						
16) L4 AR-1242-1	5.560	4.716	7066945	7858006	148.347	134.992
17) L4 AR-1242-2	5.588	4.734	8248278	5308380	120.612	66.506 #
18) L4 AR-1242-3	5.646	4.915	6600794	7273936	151.601	155.101
19) L4 AR-1242-4	5.743	4.993	6299646	3899682	178.779	84.044 #
20) L4 AR-1242-5	6.489	5.525	918133	884662	28.336m	15.823 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060123\
Data File : PP058171.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Jun 2023 06:59
Operator : YP\AJ
Sample : 02999-02
Misc :
ALS Vial : 47 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
50010

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/02/2023
Supervised By :Ankita Jodhani 06/02/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 02 07:16:29 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050823.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 08 19:18:05 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

