

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060123\
 Data File : PP058131.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Jun 2023 17:51
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 01 21:22:15 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 08 19:18:05 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.380	3.619	100.2E6	109.6E6	45.656	48.333
2) SA Decachlor...	10.202	8.676	61839896	68719124	54.569	43.777
Target Compounds						
26) L6 AR-1254-1	6.422	5.525	33908428	50515282	505.207	452.405
27) L6 AR-1254-2	6.643	5.673	49353177	44721803	521.713	448.098
28) L6 AR-1254-3	7.013	6.081	49123085	69288394	539.290	442.871
29) L6 AR-1254-4	7.301	6.311	31062401	39195492	521.513	424.874
30) L6 AR-1254-5	7.724	6.732	37177551	61746220	541.537	426.542



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060123\
Data File : PP058131.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Jun 2023 17:51
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 01 21:22:15 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050823.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 08 19:18:05 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.380 min
Delta R.T.: -0.010 min
Response: 100163036
Conc: 45.66 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.619 min
Delta R.T.: -0.004 min
Response: 109634850
Conc: 48.33 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.202 min
Delta R.T.: -0.021 min
Response: 61839896
Conc: 54.57 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.676 min
Delta R.T.: -0.010 min
Response: 68719124
Conc: 43.78 ng/ml



#26 AR-1254-1


R.T.: 6.422 min
Delta R.T.: -0.011 min
Response: 33908428
Conc: 505.21 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.525 min
Delta R.T.: -0.007 min
Response: 50515282
Conc: 452.41 ng/ml




#27 AR-1254-2

R.T.: 6.643 min
Delta R.T.: -0.011 min
Response: 49353177
Conc: 521.71 ng/ml



#27 AR-1254-2


R.T.: 5.673 min
Delta R.T.: -0.007 min
Response: 44721803
Conc: 448.10 ng/ml



#28 AR-1254-3


R.T.: 7.013 min
Delta R.T.: -0.010 min
Response: 49123085
Conc: 539.29 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.081 min
Delta R.T.: -0.007 min
Response: 69288394
Conc: 442.87 ng/ml




#29 AR-1254-4

R.T.: 7.301 min
Delta R.T.: -0.011 min
Response: 31062401
Conc: 521.51 ng/ml



#29 AR-1254-4


R.T.: 6.311 min
Delta R.T.: -0.007 min
Response: 39195492
Conc: 424.87 ng/ml



#30 AR-1254-5

R.T.: 7.724 min
Delta R.T.: -0.012 min
Response: 37177551
Conc: 541.54 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.732 min
Delta R.T.: -0.008 min
Response: 61746220
Conc: 426.54 ng/ml