

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060223\
 Data File : PP058189.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Jun 2023 13:39
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 03 05:13:31 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 08 19:18:05 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.381	3.620	103.5E6	112.7E6	47.178	49.665
2) SA Decachlor...	10.207	8.679	63368085	72423391	55.917	46.137
Target Compounds						
26) L6 AR-1254-1	6.425	5.526	34688034	54273973	516.823	486.067
27) L6 AR-1254-2	6.645	5.675	51621087	47924249	545.687	480.186
28) L6 AR-1254-3	7.015	6.082	51740486	73532334	568.024	469.997
29) L6 AR-1254-4	7.303	6.312	32337632	41453023	542.923	449.346
30) L6 AR-1254-5	7.726	6.734	39137490	65511095	570.086	452.550



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060223\
Data File : PP058189.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Jun 2023 13:39
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 03 05:13:31 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050823.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 08 19:18:05 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.381 min
Delta R.T.: -0.009 min
Response: 103501097
Conc: 47.18 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.620 min
Delta R.T.: -0.004 min
Response: 112656880
Conc: 49.67 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.207 min
Delta R.T.: -0.016 min
Response: 63368085
Conc: 55.92 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.679 min
Delta R.T.: -0.007 min
Response: 72423391
Conc: 46.14 ng/ml



#26 AR-1254-1


R.T.: 6.425 min
Delta R.T.: -0.009 min
Response: 34688034
Conc: 516.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.526 min
Delta R.T.: -0.006 min
Response: 54273973
Conc: 486.07 ng/ml




#27 AR-1254-2

R.T.: 6.645 min
Delta R.T.: -0.009 min
Response: 51621087
Conc: 545.69 ng/ml



#27 AR-1254-2


R.T.: 5.675 min
Delta R.T.: -0.006 min
Response: 47924249
Conc: 480.19 ng/ml



#28 AR-1254-3


R.T.: 7.015 min
Delta R.T.: -0.008 min
Response: 51740486
Conc: 568.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.082 min
Delta R.T.: -0.006 min
Response: 73532334
Conc: 470.00 ng/ml




#29 AR-1254-4

R.T.: 7.303 min
Delta R.T.: -0.009 min
Response: 32337632
Conc: 542.92 ng/ml



#29 AR-1254-4


R.T.: 6.312 min
Delta R.T.: -0.006 min
Response: 41453023
Conc: 449.35 ng/ml



#30 AR-1254-5

R.T.: 7.726 min
Delta R.T.: -0.010 min
Response: 39137490
Conc: 570.09 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.734 min
Delta R.T.: -0.006 min
Response: 65511095
Conc: 452.55 ng/ml