

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060225\
 Data File : PP072535.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Jun 2025 22:00
 Operator : YP\AJ
 Sample : Q2166-22
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 22

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/03/2025
 Supervised By :mohammad ahmed 06/04/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 03 03:28:10 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 24 03:32:20 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.502	3.794	38438421	33188667	19.116	20.764
2) SA Decachlor...	10.203	8.810	31202323	22973892	19.153	21.702m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060225\
Data File : PP072535.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Jun 2025 22:00
Operator : YP\AJ
Sample : Q2166-22
Misc :
ALS Vial : 30 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
22


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/03/2025
Supervised By :mohammad ahmed 06/04/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 03 03:28:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.502 min
Delta R.T.: 0.004 min
Response: 38438421
Conc: 19.12 ng/ml

Instrument :
ECD_P
ClientSampleId :
22


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/03/2025
Supervised By :mohammad ahmed 06/04/2025




#1 Tetrachloro-m-xylene

R.T.: 3.794 min
Delta R.T.: 0.000 min
Response: 33188667
Conc: 20.76 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.203 min
Delta R.T.: -0.009 min
Response: 31202323
Conc: 19.15 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.810 min
Delta R.T.: -0.011 min
Response: 22973892
Conc: 21.70 ng/ml m