

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060322\
 Data File : PP048418.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Jun 2022 20:16
 Operator : YP\AJ
 Sample : N3139-02
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 COMP-2

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/06/2022
 Supervised By :Ankita Jodhani 06/06/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 04 01:03:55 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP052722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 27 23:53:39 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.033	3.213	45858884	37758724	19.583	19.159
2) SA Decachlor...	10.385	8.638	28275624	22941164	14.548	14.192m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060322\
Data File : PP048418.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Jun 2022 20:16
Operator : YP\AJ
Sample : N3139-02
Misc :
ALS Vial : 40 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
COMP-2


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/06/2022
Supervised By :Ankita Jodhani 06/06/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 04 01:03:55 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP052722.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 27 23:53:39 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.033 min
Delta R.T.: -0.004 min
Response: 45858884
Conc: 19.58 ng/ml

Instrument :
ECD_P
ClientSampleId :
COMP-2


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/06/2022
Supervised By :Ankita Jodhani 06/06/2022




#1 Tetrachloro-m-xylene

R.T.: 3.213 min
Delta R.T.: -0.010 min
Response: 37758724
Conc: 19.16 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.385 min
Delta R.T.: -0.008 min
Response: 28275624
Conc: 14.55 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.638 min
Delta R.T.: -0.022 min
Response: 22941164
Conc: 14.19 ng/ml m