

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060324\
 Data File : PP065405.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Jun 2024 18:18
 Operator : YP\AJ
 Sample : AR1248ICC250
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 04 01:54:25 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060324.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jun 04 01:51:36 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.471	3.752	43133664	38642145	25.814	25.757
2) SA Decachlor...	10.210	8.816	46230821	40190129	27.093	26.156
Target Compounds						
21) L5 AR-1248-1	5.634	4.849	9788256	8440766	265.884	270.401
22) L5 AR-1248-2	5.907	5.088	15450598	12135131	268.233	271.677
23) L5 AR-1248-3	6.112	5.130	16872269	12736232	267.740	270.958
24) L5 AR-1248-4	6.514	5.304	17600852	14900434	266.822	270.601
25) L5 AR-1248-5	6.553	5.700	17209379	13590345	266.461	267.928



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060324\
Data File : PP065405.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Jun 2024 18:18
Operator : YP\AJ
Sample : AR1248ICC250
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 04 01:54:25 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060324.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jun 04 01:51:36 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.471 min
Delta R.T.: -0.002 min
Response: 43133664
Conc: 25.81 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.752 min
Delta R.T.: 0.000 min
Response: 38642145
Conc: 25.76 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.210 min
Delta R.T.: -0.002 min
Response: 46230821
Conc: 27.09 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.816 min
Delta R.T.: 0.000 min
Response: 40190129
Conc: 26.16 ng/ml



#21 AR-1248-1


R.T.: 5.634 min
Delta R.T.: -0.001 min
Response: 9788256
Conc: 265.88 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250




#21 AR-1248-1

R.T.: 4.849 min
Delta R.T.: 0.000 min
Response: 8440766
Conc: 270.40 ng/ml




#22 AR-1248-2

R.T.: 5.907 min
Delta R.T.: -0.001 min
Response: 15450598
Conc: 268.23 ng/ml



#22 AR-1248-2


R.T.: 5.088 min
Delta R.T.: 0.000 min
Response: 12135131
Conc: 271.68 ng/ml



#23 AR-1248-3


R.T.: 6.112 min
Delta R.T.: -0.001 min
Response: 16872269
Conc: 267.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250




#23 AR-1248-3

R.T.: 5.130 min
Delta R.T.: -0.001 min
Response: 12736232
Conc: 270.96 ng/ml




#24 AR-1248-4

R.T.: 6.514 min
Delta R.T.: -0.001 min
Response: 17600852
Conc: 266.82 ng/ml



#24 AR-1248-4


R.T.: 5.304 min
Delta R.T.: -0.001 min
Response: 14900434
Conc: 270.60 ng/ml



#25 AR-1248-5

R.T.: 6.553 min
Delta R.T.: -0.002 min
Response: 17209379
Conc: 266.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250



#25 AR-1248-5

R.T.: 5.700 min
Delta R.T.: 0.000 min
Response: 13590345
Conc: 267.93 ng/ml