

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060325\
 Data File : PP072564.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Jun 2025 15:05
 Operator : YP\AJ
 Sample : Q2176-01
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 TP-46

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 03 17:04:24 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 24 03:32:20 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.502	3.794	36007573	32341457	17.907	20.234
2)	SA Decachlor...	10.202	8.810	26299169	19671717	16.143	18.583
Target Compounds							
3)	L1 AR-1016-1	5.654	4.876	2143738	2237368	28.551	36.422 #
4)	L1 AR-1016-2	5.676	4.894	3234139	3177496	30.228	36.195
5)	L1 AR-1016-3	5.738	5.070	2246039	1692602	34.257	35.518
6)	L1 AR-1016-4	5.836	5.112	1790575	1123598	33.613	29.533
7)	L1 AR-1016-5	6.127	5.326	1587271	1693540	32.453	33.990
8)	L2 AR-1221-1	4.704	4.001	858541	334505	32.841	12.477 #
9)	L2 AR-1221-2	4.804	4.093	1045743	1663190	51.264	80.993 #
10)	L2 AR-1221-3	4.866	4.166	1245298	1088874	20.584	18.491
11)	L3 AR-1232-1	4.866	4.166	1245298	1088874	26.435	25.121
12)	L3 AR-1232-2	5.389	4.894	1637356	3177496	70.324	71.920
13)	L3 AR-1232-3	5.676	5.070	3234139	1692602	64.601	72.386
14)	L3 AR-1232-4	5.836	5.154	1790575	1262892	71.340	60.950
15)	L3 AR-1232-5	5.927	5.326	1228261	1693540	70.108	73.141
16)	L4 AR-1242-1	5.654	4.876	2143738	2237368	34.246	42.937 #
17)	L4 AR-1242-2	5.676	4.894	3234139	3177496	37.268	43.233
18)	L4 AR-1242-3	5.738	5.070	2246039	1692602	40.788	43.261
19)	L4 AR-1242-4	5.836	5.154	1790575	1262892	40.375	33.603
20)	L4 AR-1242-5	6.567	5.677	1655531	1245906	32.471	25.818
21)	L5 AR-1248-1	5.654	4.876	2143738	2237368	44.296	51.646
22)	L5 AR-1248-2	5.927	5.112	1228261	1123598	20.255	19.895
23)	L5 AR-1248-3	6.127	5.154	1587271	1262892	23.984	21.366
24)	L5 AR-1248-4	6.527	5.326	1525100	1693540	17.451	24.314 #
25)	L5 AR-1248-5	6.567	5.719	1655531	858726	19.940	12.313 #
26)	L6 AR-1254-1	6.499	5.677	1751634	1245906	20.367	12.465 #
27)	L6 AR-1254-2	6.719	5.839	1366061	122610	10.454	1.421 #
28)	L6 AR-1254-3	7.088	6.231	608639	220199	4.611	1.712 #
29)	L6 AR-1254-4	7.367	6.469	509632	288908	3.905	3.387
30)	L6 AR-1254-5	7.832f	6.908	446376	109154	4.012	0.963 #
31)	L7 AR-1260-1	7.242	6.384	903998	664599	9.659	8.333
32)	L7 AR-1260-2	7.534	6.544	788366	303443	5.494	2.992 #
33)	L7 AR-1260-3	7.859	6.764f	220275	1342352	1.934	15.447 #
34)	L7 AR-1260-4	8.064	7.168	1163476	59550	10.843	0.826 #
35)	L7 AR-1260-5	8.396	7.415	82574	1256707	0.349	7.238 #
36)	L8 AR-1262-1	8.064	6.908	1163476	109154	8.732	0.966 #
37)	L8 AR-1262-2	8.396	7.168	82574	59550	0.303	0.613 #
38)	L8 AR-1262-3	8.684f	7.702	154470	484997	0.833	5.949 #
39)	L8 AR-1262-4	8.829	7.762	3982961	55437	29.661	0.394 #
40)	L8 AR-1262-5	9.456	8.266	326493	68608	3.531	1.054 #
41)	L9 AR-1268-1	8.684	7.702	154470	484997	0.464	2.071 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060325\
Data File : PP072564.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Jun 2025 15:05
Operator : YP\AJ
Sample : Q2176-01
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
TP-46

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 03 17:04:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	8.829	7.762	3982961	55437	14.225	0.268 #
43)	L9 AR-1268-3	9.085f	7.958	574004	42971	2.377	0.254 #
44)	L9 AR-1268-4	9.456	8.266	326493	68608	3.111	0.940 #
45)	L9 AR-1268-5	9.877	8.573	294614	207447	0.435	0.454



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060325\
Data File : PP072564.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Jun 2025 15:05
Operator : YP\AJ
Sample : Q2176-01
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
TP-46

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 03 17:04:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.502 min
Delta R.T.: 0.004 min
Response: 36007573
Conc: 17.91 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-46




#1 Tetrachloro-m-xylene

R.T.: 3.794 min
Delta R.T.: 0.000 min
Response: 32341457
Conc: 20.23 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.202 min
Delta R.T.: -0.010 min
Response: 26299169
Conc: 16.14 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.810 min
Delta R.T.: -0.011 min
Response: 19671717
Conc: 18.58 ng/ml



#3 AR-1016-1


R.T.: 5.654 min
Delta R.T.: 0.004 min
Response: 2143738
Conc: 28.55 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-46




#3 AR-1016-1

R.T.: 4.876 min
Delta R.T.: -0.002 min
Response: 2237368
Conc: 36.42 ng/ml




#4 AR-1016-2

R.T.: 5.676 min
Delta R.T.: 0.004 min
Response: 3234139
Conc: 30.23 ng/ml



#4 AR-1016-2


R.T.: 4.894 min
Delta R.T.: -0.002 min
Response: 3177496
Conc: 36.19 ng/ml



#5 AR-1016-3


R.T.: 5.738 min
Delta R.T.: 0.005 min
Response: 2246039
Conc: 34.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-46




#5 AR-1016-3

R.T.: 5.070 min
Delta R.T.: -0.003 min
Response: 1692602
Conc: 35.52 ng/ml




#6 AR-1016-4

R.T.: 5.836 min
Delta R.T.: 0.004 min
Response: 1790575
Conc: 33.61 ng/ml



#6 AR-1016-4


R.T.: 5.112 min
Delta R.T.: -0.003 min
Response: 1123598
Conc: 29.53 ng/ml



#7 AR-1016-5


R.T.: 6.127 min
Delta R.T.: 0.000 min
Response: 1587271
Conc: 32.45 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-46




#7 AR-1016-5

R.T.: 5.326 min
Delta R.T.: -0.005 min
Response: 1693540
Conc: 33.99 ng/ml




#8 AR-1221-1

R.T.: 4.704 min
Delta R.T.: 0.002 min
Response: 858541
Conc: 32.84 ng/ml



#8 AR-1221-1


R.T.: 4.001 min
Delta R.T.: -0.006 min
Response: 334505
Conc: 12.48 ng/ml



#9 AR-1221-2


R.T.: 4.804 min
Delta R.T.: 0.015 min
Response: 1045743
Conc: 51.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-46




#9 AR-1221-2

R.T.: 4.093 min
Delta R.T.: 0.000 min
Response: 1663190
Conc: 80.99 ng/ml




#10 AR-1221-3

R.T.: 4.866 min
Delta R.T.: 0.003 min
Response: 1245298
Conc: 20.58 ng/ml



#10 AR-1221-3


R.T.: 4.166 min
Delta R.T.: -0.003 min
Response: 1088874
Conc: 18.49 ng/ml



#11 AR-1232-1


R.T.: 4.866 min
Delta R.T.: 0.003 min
Response: 1245298
Conc: 26.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-46




#11 AR-1232-1

R.T.: 4.166 min
Delta R.T.: -0.002 min
Response: 1088874
Conc: 25.12 ng/ml




#12 AR-1232-2

R.T.: 5.389 min
Delta R.T.: 0.000 min
Response: 1637356
Conc: 70.32 ng/ml



#12 AR-1232-2


R.T.: 4.894 min
Delta R.T.: -0.004 min
Response: 3177496
Conc: 71.92 ng/ml



#13 AR-1232-3


R.T.: 5.676 min
Delta R.T.: 0.000 min
Response: 3234139
Conc: 64.60 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-46




#13 AR-1232-3

R.T.: 5.070 min
Delta R.T.: -0.004 min
Response: 1692602
Conc: 72.39 ng/ml




#14 AR-1232-4

R.T.: 5.836 min
Delta R.T.: 0.000 min
Response: 1790575
Conc: 71.34 ng/ml



#14 AR-1232-4


R.T.: 5.154 min
Delta R.T.: -0.005 min
Response: 1262892
Conc: 60.95 ng/ml



#15 AR-1232-5


R.T.: 5.927 min
Delta R.T.: 0.002 min
Response: 1228261
Conc: 70.11 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-46




#15 AR-1232-5

R.T.: 5.326 min
Delta R.T.: -0.005 min
Response: 1693540
Conc: 73.14 ng/ml




#16 AR-1242-1

R.T.: 5.654 min
Delta R.T.: -0.002 min
Response: 2143738
Conc: 34.25 ng/ml



#16 AR-1242-1


R.T.: 4.876 min
Delta R.T.: -0.005 min
Response: 2237368
Conc: 42.94 ng/ml



#17 AR-1242-2


R.T.: 5.676 min
Delta R.T.: -0.002 min
Response: 3234139
Conc: 37.27 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-46




#17 AR-1242-2

R.T.: 4.894 min
Delta R.T.: -0.004 min
Response: 3177496
Conc: 43.23 ng/ml




#18 AR-1242-3

R.T.: 5.738 min
Delta R.T.: -0.002 min
Response: 2246039
Conc: 40.79 ng/ml



#18 AR-1242-3


R.T.: 5.070 min
Delta R.T.: -0.006 min
Response: 1692602
Conc: 43.26 ng/ml



#19 AR-1242-4


R.T.: 5.836 min
Delta R.T.: -0.002 min
Response: 1790575
Conc: 40.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-46




#19 AR-1242-4

R.T.: 5.154 min
Delta R.T.: -0.006 min
Response: 1262892
Conc: 33.60 ng/ml




#20 AR-1242-5

R.T.: 6.567 min
Delta R.T.: -0.001 min
Response: 1655531
Conc: 32.47 ng/ml



#20 AR-1242-5


R.T.: 5.677 min
Delta R.T.: -0.007 min
Response: 1245906
Conc: 25.82 ng/ml



#21 AR-1248-1


R.T.: 5.654 min
Delta R.T.: -0.004 min
Response: 2143738
Conc: 44.30 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-46




#21 AR-1248-1

R.T.: 4.876 min
Delta R.T.: -0.005 min
Response: 2237368
Conc: 51.65 ng/ml




#22 AR-1248-2

R.T.: 5.927 min
Delta R.T.: -0.004 min
Response: 1228261
Conc: 20.25 ng/ml



#22 AR-1248-2


R.T.: 5.112 min
Delta R.T.: -0.006 min
Response: 1123598
Conc: 19.89 ng/ml



#23 AR-1248-3


R.T.: 6.127 min
Delta R.T.: -0.007 min
Response: 1587271
Conc: 23.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-46




#23 AR-1248-3

R.T.: 5.154 min
Delta R.T.: -0.007 min
Response: 1262892
Conc: 21.37 ng/ml




#24 AR-1248-4

R.T.: 6.527 min
Delta R.T.: -0.006 min
Response: 1525100
Conc: 17.45 ng/ml



#24 AR-1248-4


R.T.: 5.326 min
Delta R.T.: -0.007 min
Response: 1693540
Conc: 24.31 ng/ml



#25 AR-1248-5


R.T.: 6.567 min
Delta R.T.: -0.005 min
Response: 1655531
Conc: 19.94 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-46




#25 AR-1248-5

R.T.: 5.719 min
Delta R.T.: -0.006 min
Response: 858726
Conc: 12.31 ng/ml




#26 AR-1254-1

R.T.: 6.499 min
Delta R.T.: -0.008 min
Response: 1751634
Conc: 20.37 ng/ml



#26 AR-1254-1


R.T.: 5.677 min
Delta R.T.: -0.008 min
Response: 1245906
Conc: 12.46 ng/ml



#27 AR-1254-2


R.T.: 6.719 min
Delta R.T.: -0.005 min
Response: 1366061
Conc: 10.45 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-46




#27 AR-1254-2

R.T.: 5.839 min
Delta R.T.: 0.005 min
Response: 122610
Conc: 1.42 ng/ml




#28 AR-1254-3

R.T.: 7.088 min
Delta R.T.: 0.000 min
Response: 608639
Conc: 4.61 ng/ml



#28 AR-1254-3


R.T.: 6.231 min
Delta R.T.: -0.005 min
Response: 220199
Conc: 1.71 ng/ml



#29 AR-1254-4


R.T.: 7.367 min
Delta R.T.: -0.003 min
Response: 509632
Conc: 3.90 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-46




#29 AR-1254-4

R.T.: 6.469 min
Delta R.T.: 0.004 min
Response: 288908
Conc: 3.39 ng/ml




#30 AR-1254-5

R.T.: 7.832 min
Delta R.T.: 0.045 min
Response: 446376
Conc: 4.01 ng/ml



#30 AR-1254-5


R.T.: 6.908 min
Delta R.T.: 0.025 min
Response: 109154
Conc: 0.96 ng/ml



#31 AR-1260-1


R.T.: 7.242 min
Delta R.T.: -0.005 min
Response: 903998
Conc: 9.66 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-46




#31 AR-1260-1

R.T.: 6.384 min
Delta R.T.: 0.021 min
Response: 664599
Conc: 8.33 ng/ml




#32 AR-1260-2

R.T.: 7.534 min
Delta R.T.: 0.033 min
Response: 788366
Conc: 5.49 ng/ml



#32 AR-1260-2


R.T.: 6.544 min
Delta R.T.: -0.008 min
Response: 303443
Conc: 2.99 ng/ml



#33 AR-1260-3


R.T.: 7.859 min
Delta R.T.: 0.000 min
Response: 220275
Conc: 1.93 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-46




#33 AR-1260-3

R.T.: 6.764 min
Delta R.T.: 0.059 min
Response: 1342352
Conc: 15.45 ng/ml




#34 AR-1260-4

R.T.: 8.064 min
Delta R.T.: -0.019 min
Response: 1163476
Conc: 10.84 ng/ml



#34 AR-1260-4


R.T.: 7.168 min
Delta R.T.: -0.008 min
Response: 59550
Conc: 0.83 ng/ml



#35 AR-1260-5


R.T.: 8.396 min
Delta R.T.: -0.007 min
Response: 82574
Conc: 0.35 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-46




#35 AR-1260-5

R.T.: 7.415 min
Delta R.T.: -0.003 min
Response: 1256707
Conc: 7.24 ng/ml




#36 AR-1262-1

R.T.: 8.064 min
Delta R.T.: -0.024 min
Response: 1163476
Conc: 8.73 ng/ml



#36 AR-1262-1


R.T.: 6.908 min
Delta R.T.: -0.013 min
Response: 109154
Conc: 0.97 ng/ml



#37 AR-1262-2


R.T.: 8.396 min
Delta R.T.: -0.011 min
Response: 82574
Conc: 0.30 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-46




#37 AR-1262-2

R.T.: 7.168 min
Delta R.T.: -0.011 min
Response: 59550
Conc: 0.61 ng/ml




#38 AR-1262-3

R.T.: 8.684 min
Delta R.T.: -0.039 min
Response: 154470
Conc: 0.83 ng/ml



#38 AR-1262-3


R.T.: 7.702 min
Delta R.T.: 0.000 min
Response: 484997
Conc: 5.95 ng/ml



#39 AR-1262-4


R.T.: 8.829 min
Delta R.T.: 0.020 min
Response: 3982961
Conc: 29.66 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-46




#39 AR-1262-4

R.T.: 7.762 min
Delta R.T.: -0.005 min
Response: 55437
Conc: 0.39 ng/ml




#40 AR-1262-5

R.T.: 9.456 min
Delta R.T.: -0.008 min
Response: 326493
Conc: 3.53 ng/ml



#40 AR-1262-5


R.T.: 8.266 min
Delta R.T.: -0.002 min
Response: 68608
Conc: 1.05 ng/ml



#41 AR-1268-1


R.T.: 8.684 min
Delta R.T.: -0.033 min
Response: 154470
Conc: 0.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-46




#41 AR-1268-1

R.T.: 7.702 min
Delta R.T.: 0.000 min
Response: 484997
Conc: 2.07 ng/ml




#42 AR-1268-2

R.T.: 8.829 min
Delta R.T.: 0.018 min
Response: 3982961
Conc: 14.23 ng/ml



#42 AR-1268-2


R.T.: 7.762 min
Delta R.T.: -0.006 min
Response: 55437
Conc: 0.27 ng/ml



#43 AR-1268-3


R.T.: 9.085 min
Delta R.T.: 0.045 min
Response: 574004
Conc: 2.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-46




#43 AR-1268-3

R.T.: 7.958 min
Delta R.T.: -0.014 min
Response: 42971
Conc: 0.25 ng/ml




#44 AR-1268-4

R.T.: 9.456 min
Delta R.T.: -0.004 min
Response: 326493
Conc: 3.11 ng/ml



#44 AR-1268-4


R.T.: 8.266 min
Delta R.T.: 0.000 min
Response: 68608
Conc: 0.94 ng/ml



#45 AR-1268-5

R.T.: 9.877 min
Delta R.T.: 0.001 min
Response: 294614
Conc: 0.44 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-46



#45 AR-1268-5

R.T.: 8.573 min
Delta R.T.: 0.008 min
Response: 207447
Conc: 0.45 ng/ml