

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060524\
 Data File : PP065501.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Jun 2024 21:53
 Operator : YP\AJ
 Sample : AR1248ICC1000
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 06 05:06:00 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060524.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 06 05:05:02 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.826	4.071	181.6E6	267.1E6	98.066	95.124
2) SA Decachlor...	10.842	9.259	178.5E6	137.5E6	92.623	92.285
Target Compounds						
21) L5 AR-1248-1	6.004	5.186	37916601	49799034	938.623	898.593
22) L5 AR-1248-2	6.278	5.425	57014905	65934085	925.549	895.649
23) L5 AR-1248-3	6.486	5.469	61557469	68813060	914.345	898.640
24) L5 AR-1248-4	6.890	5.646	65310832	79729007	901.557	905.508
25) L5 AR-1248-5	6.930	6.044	63970925	72610787	905.562	913.629



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060524\
Data File : PP065501.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Jun 2024 21:53
Operator : YP\AJ
Sample : AR1248ICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 06 05:06:00 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060524.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 06 05:05:02 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.826 min
Delta R.T.: 0.000 min
Response: 181553080
Conc: 98.07 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000




#1 Tetrachloro-m-xylene

R.T.: 4.071 min
Delta R.T.: 0.000 min
Response: 267053358
Conc: 95.12 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.842 min
Delta R.T.: 0.000 min
Response: 178489277
Conc: 92.62 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.259 min
Delta R.T.: 0.000 min
Response: 137458807
Conc: 92.29 ng/ml



#21 AR-1248-1


R.T.: 6.004 min
Delta R.T.: 0.000 min
Response: 37916601
Conc: 938.62 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000




#21 AR-1248-1

R.T.: 5.186 min
Delta R.T.: 0.000 min
Response: 49799034
Conc: 898.59 ng/ml




#22 AR-1248-2

R.T.: 6.278 min
Delta R.T.: 0.000 min
Response: 57014905
Conc: 925.55 ng/ml



#22 AR-1248-2


R.T.: 5.425 min
Delta R.T.: 0.000 min
Response: 65934085
Conc: 895.65 ng/ml



#23 AR-1248-3


R.T.: 6.486 min
Delta R.T.: 0.000 min
Response: 61557469
Conc: 914.35 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000




#23 AR-1248-3

R.T.: 5.469 min
Delta R.T.: 0.000 min
Response: 68813060
Conc: 898.64 ng/ml




#24 AR-1248-4

R.T.: 6.890 min
Delta R.T.: 0.000 min
Response: 65310832
Conc: 901.56 ng/ml



#24 AR-1248-4


R.T.: 5.646 min
Delta R.T.: 0.000 min
Response: 79729007
Conc: 905.51 ng/ml



#25 AR-1248-5

R.T.: 6.930 min
Delta R.T.: 0.000 min
Response: 63970925
Conc: 905.56 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000



#25 AR-1248-5

R.T.: 6.044 min
Delta R.T.: 0.000 min
Response: 72610787
Conc: 913.63 ng/ml