

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060524\
 Data File : PP065509.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Jun 2024 00:04
 Operator : YP\AJ
 Sample : AR1254ICC250
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC250

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/06/2024
 Supervised By :Ankita Jodhani 06/06/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 06 04:59:13 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060524.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 06 04:45:11 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.841	4.071	49196676	72504179	25.602	25.760
2) SA Decachlor...	10.858	9.260	52377457	40612844	27.212	27.102
Target Compounds						
26) L6 AR-1254-1	6.880	6.003	20496766	32815428	266.725	271.243
27) L6 AR-1254-2	7.098	6.150	31364468	28129314	273.215	273.308
28) L6 AR-1254-3	7.467	6.562	32310026	41989857	270.307	268.509
29) L6 AR-1254-4	7.751	6.790	23761100	25009542	273.173	273.455m
30) L6 AR-1254-5	8.173	7.213	26144299	34528955	265.808m	271.403

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060524\
Data File : PP065509.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Jun 2024 00:04
Operator : YP\AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/06/2024
Supervised By :Ankita Jodhani 06/06/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 06 04:59:13 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060524.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 06 04:45:11 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm





#1 Tetrachloro-m-xylene

R.T.: 4.841 min
Delta R.T.: 0.000 min
Response: 49196676
Conc: 25.60 ng/ml

Instrument :

ECD_P


ClientSampleId :

AR1254ICC250

Manual Integrations


APPROVED

Reviewed By :Yogesh Patel 06/06/2024
Supervised By :Ankita Jodhani 06/06/2024




#1 Tetrachloro-m-xylene

R.T.: 4.071 min
Delta R.T.: 0.000 min
Response: 72504179
Conc: 25.76 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.858 min
Delta R.T.: 0.000 min
Response: 52377457
Conc: 27.21 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.260 min
Delta R.T.: 0.000 min
Response: 40612844
Conc: 27.10 ng/ml




#26 AR-1254-1

R.T.: 6.880 min
Delta R.T.: 0.000 min
Response: 20496766
Conc: 266.72 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/06/2024
Supervised By :Ankita Jodhani 06/06/2024




#26 AR-1254-1

R.T.: 6.003 min
Delta R.T.: 0.000 min
Response: 32815428
Conc: 271.24 ng/ml




#27 AR-1254-2

R.T.: 7.098 min
Delta R.T.: 0.000 min
Response: 31364468
Conc: 273.22 ng/ml



#27 AR-1254-2

R.T.: 6.150 min
Delta R.T.: 0.000 min
Response: 28129314
Conc: 273.31 ng/ml




#28 AR-1254-3

R.T.: 7.467 min
Delta R.T.: 0.000 min
Response: 32310026
Conc: 270.31 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/06/2024
Supervised By :Ankita Jodhani 06/06/2024




#28 AR-1254-3

R.T.: 6.562 min
Delta R.T.: 0.000 min
Response: 41989857
Conc: 268.51 ng/ml




#29 AR-1254-4

R.T.: 7.751 min
Delta R.T.: 0.000 min
Response: 23761100
Conc: 273.17 ng/ml



#29 AR-1254-4

R.T.: 6.790 min
Delta R.T.: 0.000 min
Response: 25009542
Conc: 273.46 ng/ml m




#30 AR-1254-5

R.T.: 8.173 min
Delta R.T.: 0.000 min
Response: 26144299
Conc: 265.81 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/06/2024
Supervised By :Ankita Jodhani 06/06/2024



#30 AR-1254-5

R.T.: 7.213 min
Delta R.T.: 0.000 min
Response: 34528955
Conc: 271.40 ng/ml