

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060622\
 Data File : PP048455.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Jun 2022 10:10
 Operator : YP\AJ
 Sample : N2988-10DL 2X
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 P015-SS003-0612-01DL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 06 15:09:21 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP052722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 27 23:53:39 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.034	3.213	25667932	20329377	10.961	10.315
2) SA Decachlor...	10.381	8.636	19835822	14667137	10.206	9.073
Target Compounds						
31) L7 AR-1260-1	7.091	5.972	72757882	76025726	779.524	888.115
32) L7 AR-1260-2	7.376	6.178	104.2E6	64221498	932.335	623.041 #
33) L7 AR-1260-3	7.775	6.344	63152672	62075810	737.276	626.886
34) L7 AR-1260-4	8.027	6.860	74704536	57322866	719.996	688.558
35) L7 AR-1260-5	8.389	7.125	141.1E6	134.4E6	683.588	677.344

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060622\
Data File : PP048455.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Jun 2022 10:10
Operator : YP\AJ
Sample : N2988-10DL 2X
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
P015-SS003-0612-01DL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 06 15:09:21 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP052722.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 27 23:53:39 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

