

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060721\
 Data File : PP036249.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 07 Jun 2021 12:35
 Operator : DD\AJ
 Sample : M2583-19
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 33-34

Manual Integrations
 APPROVED

Ankita
 6/8/2021 3:22:09 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 07 15:15:58 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP052521.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed May 26 00:48:29 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.964	3.960	893957	587742	20.085m	21.059
2) SA Decachlor...	10.991	9.266	565952	355249	13.243	15.485

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060721\
Data File : PP036249.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 07 Jun 2021 12:35
Operator : DD\AJ
Sample : M2583-19
Misc :
ALS Vial : 13 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
33-34


Manual Integrations
APPROVED

Ankita
6/8/2021 3:22:09 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 07 15:15:58 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP052521.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed May 26 00:48:29 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





#1 Tetrachloro-m-xylene

R.T.: 4.964 min

Delta R.T.: -0.001 min


Response: 893957

Conc: 20.08 ng/ml m

Instrument :
ECD_P
ClientSampleId :
33-34

Manual Integrations
APPROVED

Ankita
6/8/2021 3:22:09 PM




#1 Tetrachloro-m-xylene

R.T.: 3.960 min

Delta R.T.: -0.004 min

Response: 587742

Conc: 21.06 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.991 min

Delta R.T.: -0.001 min

Response: 565952

Conc: 13.24 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.266 min

Delta R.T.: -0.008 min

Response: 355249

Conc: 15.48 ng/ml