

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060722\
 Data File : PP048489.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Jun 2022 01:59
 Operator : YP\AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP060722AR1254

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 07 10:08:01 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jun 07 10:06:29 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.035	3.204	126.9E6	103.1E6	51.190	50.572
2) SA Decachlor...	10.386	8.625	102.7E6	65056152	51.527	51.263
Target Compounds						
26) L6 AR-1254-1	6.264	5.226	46035336	49989915	495.007	494.511
27) L6 AR-1254-2	6.507	5.386	69537656	43667829	501.314	495.932
28) L6 AR-1254-3	6.915	5.822	71945470	69104005	506.009	497.121
29) L6 AR-1254-4	7.233	6.072	53771272	43284544	504.591	499.150
30) L6 AR-1254-5	7.698	6.526	59853431	61438927	504.192	500.948

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060722\
Data File : PP048489.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Jun 2022 01:59
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP060722AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 07 10:08:01 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060722.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jun 07 10:06:29 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

