

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060722\
 Data File : PP048483.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Jun 2022 00:27
 Operator : YP\AJ
 Sample : AR1254ICC050
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 07 10:06:37 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jun 07 10:06:29 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.035	3.204	12435326	10555555	5.015	5.179
2) SA Decachlor...	10.384	8.626	9670920	6921274	4.850	5.454
Target Compounds						
26) L6 AR-1254-1	6.264	5.226	5308354	5720854	57.079	56.592
27) L6 AR-1254-2	6.507	5.386	7562207	5018234	54.518	56.992
28) L6 AR-1254-3	6.914	5.822	7467460	7806657	52.520	56.160
29) L6 AR-1254-4	7.233	6.072	5568571	4760151	52.256	54.893
30) L6 AR-1254-5	7.698	6.526	6574135	6713644	55.379	54.740



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060722\
Data File : PP048483.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Jun 2022 00:27
Operator : YP\AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 07 10:06:37 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060722.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jun 07 10:06:29 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.035 min
Delta R.T.: 0.000 min
Response: 12435326
Conc: 5.01 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050




#1 Tetrachloro-m-xylene

R.T.: 3.204 min
Delta R.T.: 0.000 min
Response: 10555555
Conc: 5.18 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.384 min
Delta R.T.: 0.000 min
Response: 9670920
Conc: 4.85 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.626 min
Delta R.T.: 0.000 min
Response: 6921274
Conc: 5.45 ng/ml



#26 AR-1254-1


R.T.: 6.264 min
Delta R.T.: 0.000 min
Response: 5308354
Conc: 57.08 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050




#26 AR-1254-1

R.T.: 5.226 min
Delta R.T.: 0.000 min
Response: 5720854
Conc: 56.59 ng/ml




#27 AR-1254-2

R.T.: 6.507 min
Delta R.T.: 0.000 min
Response: 7562207
Conc: 54.52 ng/ml



#27 AR-1254-2


R.T.: 5.386 min
Delta R.T.: 0.000 min
Response: 5018234
Conc: 56.99 ng/ml



#28 AR-1254-3


R.T.: 6.914 min
Delta R.T.: 0.000 min
Response: 7467460
Conc: 52.52 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050




#28 AR-1254-3

R.T.: 5.822 min
Delta R.T.: 0.000 min
Response: 7806657
Conc: 56.16 ng/ml




#29 AR-1254-4

R.T.: 7.233 min
Delta R.T.: 0.000 min
Response: 5568571
Conc: 52.26 ng/ml



#29 AR-1254-4


R.T.: 6.072 min
Delta R.T.: 0.000 min
Response: 4760151
Conc: 54.89 ng/ml



#30 AR-1254-5

R.T.: 7.698 min
Delta R.T.: 0.000 min
Response: 6574135
Conc: 55.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050



#30 AR-1254-5

R.T.: 6.526 min
Delta R.T.: 0.000 min
Response: 6713644
Conc: 54.74 ng/ml