

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060821\
 Data File : PP036283.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 08 Jun 2021 8:58
 Operator : DD\AJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 09 04:05:25 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP052521.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed May 26 00:48:29 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
2)	SA Decachlor...	10.987	0.000	53104	0	1.243	N.D. #
Target Compounds							
20)	L4 AR-1242-5	7.249	6.116f	8284	33605	8.690	51.930 #
24)	L5 AR-1248-4	7.209f	0.000	96053	0	60.313	N.D. #
25)	L5 AR-1248-5	7.249	6.116f	8284	33605	5.306	38.215 #
26)	L6 AR-1254-1	7.209f	6.116f	96053	33605	59.287	25.716 #
28)	L6 AR-1254-3	7.830f	0.000	32156	0	12.189	N.D. #
29)	L6 AR-1254-4	8.090	0.000	389	0	0.193	N.D. #
38)	L8 AR-1262-3	0.000	8.168	0	54031	N.D.	45.471 #
41)	L9 AR-1268-1	0.000	8.168f	0	54031	N.D.	15.908 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060821\
Data File : PP036283.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 08 Jun 2021 8:58
Operator : DD\AJ
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 09 04:05:25 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP052521.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed May 26 00:48:29 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 0.000 min
Exp R.T. : 4.965 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T. : 3.964 min
Response: 0
Conc: N.D.




#2 Decachlorobiphenyl

R.T.: 10.987 min
Delta R.T.: -0.006 min
Response: 53104
Conc: 1.24 ng/ml



#2 Decachlorobiphenyl


R.T.: 0.000 min
Exp R.T. : 9.275 min
Response: 0
Conc: N.D.



#20 AR-1242-5


R.T.: 7.249 min
Delta R.T.: -0.014 min
Response: 8284
Conc: 8.69 ng/ml

Instrument :
ECD_P
ClientSampleId :




#20 AR-1242-5

R.T.: 6.116 min
Delta R.T.: 0.024 min
Response: 33605
Conc: 51.93 ng/ml




#24 AR-1248-4

R.T.: 7.209 min
Delta R.T.: -0.015 min
Response: 96053
Conc: 60.31 ng/ml



#24 AR-1248-4


R.T.: 0.000 min
Exp R.T.: 5.712 min
Response: 0
Conc: N.D.



#25 AR-1248-5


R.T.: 7.249 min
Delta R.T.: -0.014 min
Response: 8284
Conc: 5.31 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 6.116 min
Delta R.T.: -0.020 min
Response: 33605
Conc: 38.21 ng/ml




#26 AR-1254-1

R.T.: 7.209 min
Delta R.T.: 0.017 min
Response: 96053
Conc: 59.29 ng/ml



#26 AR-1254-1


R.T.: 6.116 min
Delta R.T.: 0.024 min
Response: 33605
Conc: 25.72 ng/ml



#28 AR-1254-3


R.T.: 7.830 min
Delta R.T.: 0.027 min
Response: 32156
Conc: 12.19 ng/ml

Instrument :
ECD_P
ClientSampleId :




#28 AR-1254-3

R.T.: 0.000 min
Exp R.T. : 6.673 min
Response: 0
Conc: N.D.




#29 AR-1254-4

R.T.: 8.090 min
Delta R.T.: -0.009 min
Response: 389
Conc: 0.19 ng/ml



#29 AR-1254-4


R.T.: 0.000 min
Exp R.T. : 6.912 min
Response: 0
Conc: N.D.



#38 AR-1262-3


R.T.: 0.000 min
Exp R.T.: 9.470 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#38 AR-1262-3

R.T.: 8.168 min
Delta R.T.: -0.015 min
Response: 54031
Conc: 45.47 ng/ml



#41 AR-1268-1

R.T.: 0.000 min
Exp R.T.: 9.470 min
Response: 0
Conc: N.D.



#41 AR-1268-1

R.T.: 8.168 min
Delta R.T.: -0.016 min
Response: 54031
Conc: 15.91 ng/ml