

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060823\
 Data File : PP058341.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Jun 2023 22:27
 Operator : YP\AJ
 Sample : 03065-06 10X
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 08 04:49:03 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jun 06 04:59:28 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.379	3.618	4372747	4628979	1.911	2.145
2)	SA Decachlor...	10.207	8.681	3548114	2947401	3.176	1.788 #
Target Compounds							
3)	L1 AR-1016-1	5.542f	4.698f	212051	20532	3.331	0.293 #
4)	L1 AR-1016-2	5.594	0.000	135738	0	1.470	N.D. #
5)	L1 AR-1016-3	5.644	4.907	51943	56249	0.885	1.001
6)	L1 AR-1016-4	5.738	4.952	124639	50011	2.617	1.118 #
7)	L1 AR-1016-5	6.061f	0.000	439761	0	8.688	N.D. #
8)	L2 AR-1221-1	4.583	0.000	43554	0	1.539	N.D. #
9)	L2 AR-1221-2	4.666	3.923	31175	93196	1.475	4.598 #
10)	L2 AR-1221-3	4.763	0.000	66929	0	1.125	N.D. #
11)	L3 AR-1232-1	4.763	0.000	66929	0	1.481	N.D. #
12)	L3 AR-1232-2	5.287	0.000	45449	0	1.622	N.D. #
13)	L3 AR-1232-3	5.594	4.907	135738	56249	3.128	2.150 #
14)	L3 AR-1232-4	5.738	4.996	124639	48173	5.484	2.038 #
15)	L3 AR-1232-5	5.806f	0.000	8511	0	0.428	N.D. #
16)	L4 AR-1242-1	5.542f	4.698f	212051	20532	4.172	0.371 #
17)	L4 AR-1242-2	5.594	0.000	135738	0	1.846	N.D. #
18)	L4 AR-1242-3	5.644	4.907	51943	56249	1.109	1.273
19)	L4 AR-1242-4	5.738	4.996	124639	48173	3.270	1.084 #
20)	L4 AR-1242-5	6.479	5.525	105403	11737	2.976	0.212 #
21)	L5 AR-1248-1	5.542f	4.698f	212051	20532	5.301	0.466 #
22)	L5 AR-1248-2	0.000	4.952	0	50011	N.D.	0.807 #
23)	L5 AR-1248-3	6.061f	4.996	439761	48173	6.388	0.735 #
24)	L5 AR-1248-4	6.449	0.000	196573	0	3.116	N.D. #
25)	L5 AR-1248-5	6.479	5.578	105403	35217	1.650	0.477 #
26)	L6 AR-1254-1	6.420	5.525	217660	11737	2.979	0.108 #
27)	L6 AR-1254-2	6.648	5.673	39313	246709	0.383	2.541 #
28)	L6 AR-1254-3	7.015	6.082	332927	43995	3.377	0.284 #
29)	L6 AR-1254-4	7.302	6.340f	62140	35707	0.936	0.381 #
30)	L6 AR-1254-5	7.730	6.731	448108	158567	6.124	1.091 #
31)	L7 AR-1260-1	7.196f	6.211	579746	62893	6.917	0.578 #
32)	L7 AR-1260-2	0.000	6.388	0	234937	N.D.	1.824 #
33)	L7 AR-1260-3	7.798	6.548	194937	205232	3.396	1.664 #
34)	L7 AR-1260-4	8.024	7.036	283915	154792	4.201	1.710 #
35)	L7 AR-1260-5	8.367	7.282	372570	50928	3.473	0.251 #
36)	L8 AR-1262-1	7.798	6.822	194937	34152	2.119	0.546 #
37)	L8 AR-1262-2	8.367	7.036	372570	154792	2.834	1.231 #
38)	L8 AR-1262-3	8.678	7.563	59269	165043	0.628	1.663 #
39)	L8 AR-1262-4	8.785f	7.625	6815	162184	0.093	0.897 #
40)	L8 AR-1262-5	9.433	0.000	9550	0	0.215	N.D. #
41)	L9 AR-1268-1	8.665	7.563	47995	165043	0.300	0.597 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060823\
Data File : PP058341.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Jun 2023 22:27
Operator : YP\AJ
Sample : 03065-06 10X
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 08 04:49:03 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060623.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jun 06 04:59:28 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	8.785f	7.625	6815	162184	0.049	0.638 #
43)	L9 AR-1268-3	8.997	7.833	194031	68349	1.502	0.320 #
44)	L9 AR-1268-4	9.433	0.000	9550	0	0.202	N.D. #
45)	L9 AR-1268-5	9.885f	8.425	90213	68054	0.271	0.118 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060823\
Data File : PP058341.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Jun 2023 22:27
Operator : YP\AJ
Sample : 03065-06 10X
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 08 04:49:03 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060623.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jun 06 04:59:28 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.379 min
Delta R.T.: 0.000 min
Response: 4372747
Conc: 1.91 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.618 min
Delta R.T.: 0.000 min
Response: 4628979
Conc: 2.15 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.207 min
Delta R.T.: 0.003 min
Response: 3548114
Conc: 3.18 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.681 min
Delta R.T.: 0.003 min
Response: 2947401
Conc: 1.79 ng/ml



#3 AR-1016-1


R.T.: 5.542 min
Delta R.T.: -0.016 min
Response: 212051
Conc: 3.33 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.698 min
Delta R.T.: -0.016 min
Response: 20532
Conc: 0.29 ng/ml




#4 AR-1016-2

R.T.: 5.594 min
Delta R.T.: 0.013 min
Response: 135738
Conc: 1.47 ng/ml



#4 AR-1016-2


R.T.: 0.000 min
Exp R.T. : 4.733 min
Response: 0
Conc: N.D.



#5 AR-1016-3


R.T.: 5.644 min
Delta R.T.: 0.000 min
Response: 51943
Conc: 0.88 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.907 min
Delta R.T.: -0.004 min
Response: 56249
Conc: 1.00 ng/ml




#6 AR-1016-4

R.T.: 5.738 min
Delta R.T.: -0.006 min
Response: 124639
Conc: 2.62 ng/ml



#6 AR-1016-4


R.T.: 4.952 min
Delta R.T.: -0.001 min
Response: 50011
Conc: 1.12 ng/ml



#7 AR-1016-5


R.T.: 6.061 min
Delta R.T.: 0.019 min
Response: 439761
Conc: 8.69 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 0.000 min
Exp R.T. : 5.169 min
Response: 0
Conc: N.D.




#8 AR-1221-1

R.T.: 4.583 min
Delta R.T.: -0.006 min
Response: 43554
Conc: 1.54 ng/ml



#8 AR-1221-1


R.T.: 0.000 min
Exp R.T. : 3.833 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 4.666 min
Delta R.T.: -0.010 min
Response: 31175
Conc: 1.48 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.923 min
Delta R.T.: 0.003 min
Response: 93196
Conc: 4.60 ng/ml




#10 AR-1221-3

R.T.: 4.763 min
Delta R.T.: 0.010 min
Response: 66929
Conc: 1.13 ng/ml



#10 AR-1221-3


R.T.: 0.000 min
Exp R.T.: 3.996 min
Response: 0
Conc: N.D.



#11 AR-1232-1


R.T.: 4.763 min
Delta R.T.: 0.010 min
Response: 66929
Conc: 1.48 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 0.000 min
Exp R.T. : 3.996 min
Response: 0
Conc: N.D.




#12 AR-1232-2

R.T.: 5.287 min
Delta R.T.: -0.001 min
Response: 45449
Conc: 1.62 ng/ml



#12 AR-1232-2


R.T.: 0.000 min
Exp R.T. : 4.733 min
Response: 0
Conc: N.D.



#13 AR-1232-3


R.T.: 5.594 min
Delta R.T.: 0.013 min
Response: 135738
Conc: 3.13 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.907 min
Delta R.T.: -0.004 min
Response: 56249
Conc: 2.15 ng/ml




#14 AR-1232-4

R.T.: 5.738 min
Delta R.T.: -0.006 min
Response: 124639
Conc: 5.48 ng/ml



#14 AR-1232-4


R.T.: 4.996 min
Delta R.T.: 0.002 min
Response: 48173
Conc: 2.04 ng/ml



#15 AR-1232-5


R.T.: 5.806 min
Delta R.T.: -0.030 min
Response: 8511
Conc: 0.43 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 0.000 min
Exp R.T. : 5.169 min
Response: 0
Conc: N.D.




#16 AR-1242-1

R.T.: 5.542 min
Delta R.T.: -0.017 min
Response: 212051
Conc: 4.17 ng/ml



#16 AR-1242-1


R.T.: 4.698 min
Delta R.T.: -0.016 min
Response: 20532
Conc: 0.37 ng/ml



#17 AR-1242-2


R.T.: 5.594 min
Delta R.T.: 0.012 min
Response: 135738
Conc: 1.85 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 0.000 min
Exp R.T. : 4.733 min
Response: 0
Conc: N.D.




#18 AR-1242-3

R.T.: 5.644 min
Delta R.T.: 0.000 min
Response: 51943
Conc: 1.11 ng/ml



#18 AR-1242-3


R.T.: 4.907 min
Delta R.T.: -0.004 min
Response: 56249
Conc: 1.27 ng/ml



#19 AR-1242-4


R.T.: 5.738 min
Delta R.T.: -0.006 min
Response: 124639
Conc: 3.27 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 4.996 min
Delta R.T.: 0.000 min
Response: 48173
Conc: 1.08 ng/ml




#20 AR-1242-5

R.T.: 6.479 min
Delta R.T.: -0.010 min
Response: 105403
Conc: 2.98 ng/ml



#20 AR-1242-5


R.T.: 5.525 min
Delta R.T.: 0.001 min
Response: 11737
Conc: 0.21 ng/ml



#21 AR-1248-1


R.T.: 5.542 min
Delta R.T.: -0.017 min
Response: 212051
Conc: 5.30 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.698 min
Delta R.T.: -0.016 min
Response: 20532
Conc: 0.47 ng/ml




#22 AR-1248-2

R.T.: 0.000 min
Exp R.T. : 5.837 min
Response: 0
Conc: N.D.



#22 AR-1248-2


R.T.: 4.952 min
Delta R.T.: -0.001 min
Response: 50011
Conc: 0.81 ng/ml



#23 AR-1248-3


R.T.: 6.061 min
Delta R.T.: 0.019 min
Response: 439761
Conc: 6.39 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 4.996 min
Delta R.T.: 0.001 min
Response: 48173
Conc: 0.73 ng/ml




#24 AR-1248-4

R.T.: 6.449 min
Delta R.T.: 0.000 min
Response: 196573
Conc: 3.12 ng/ml



#24 AR-1248-4


R.T.: 0.000 min
Exp R.T. : 5.169 min
Response: 0
Conc: N.D.



#25 AR-1248-5


R.T.: 6.479 min
Delta R.T.: -0.010 min
Response: 105403
Conc: 1.65 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.578 min
Delta R.T.: 0.012 min
Response: 35217
Conc: 0.48 ng/ml




#26 AR-1254-1

R.T.: 6.420 min
Delta R.T.: -0.004 min
Response: 217660
Conc: 2.98 ng/ml



#26 AR-1254-1


R.T.: 5.525 min
Delta R.T.: 0.000 min
Response: 11737
Conc: 0.11 ng/ml



#27 AR-1254-2


R.T.: 6.648 min
Delta R.T.: 0.004 min
Response: 39313
Conc: 0.38 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.673 min
Delta R.T.: -0.001 min
Response: 246709
Conc: 2.54 ng/ml




#28 AR-1254-3

R.T.: 7.015 min
Delta R.T.: 0.000 min
Response: 332927
Conc: 3.38 ng/ml



#28 AR-1254-3


R.T.: 6.082 min
Delta R.T.: 0.001 min
Response: 43995
Conc: 0.28 ng/ml



#29 AR-1254-4


R.T.: 7.302 min
Delta R.T.: 0.000 min
Response: 62140
Conc: 0.94 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.340 min
Delta R.T.: 0.028 min
Response: 35707
Conc: 0.38 ng/ml




#30 AR-1254-5

R.T.: 7.730 min
Delta R.T.: 0.004 min
Response: 448108
Conc: 6.12 ng/ml



#30 AR-1254-5


R.T.: 6.731 min
Delta R.T.: -0.002 min
Response: 158567
Conc: 1.09 ng/ml



#31 AR-1260-1


R.T.: 7.196 min
Delta R.T.: 0.016 min
Response: 579746
Conc: 6.92 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.211 min
Delta R.T.: 0.000 min
Response: 62893
Conc: 0.58 ng/ml




#32 AR-1260-2

R.T.: 0.000 min
Exp R.T. : 7.439 min
Response: 0
Conc: N.D.



#32 AR-1260-2


R.T.: 6.388 min
Delta R.T.: -0.013 min
Response: 234937
Conc: 1.82 ng/ml



#33 AR-1260-3


R.T.: 7.798 min
Delta R.T.: -0.004 min
Response: 194937
Conc: 3.40 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.548 min
Delta R.T.: -0.009 min
Response: 205232
Conc: 1.66 ng/ml




#34 AR-1260-4

R.T.: 8.024 min
Delta R.T.: -0.007 min
Response: 283915
Conc: 4.20 ng/ml



#34 AR-1260-4


R.T.: 7.036 min
Delta R.T.: 0.004 min
Response: 154792
Conc: 1.71 ng/ml



#35 AR-1260-5


R.T.: 8.367 min
Delta R.T.: 0.014 min
Response: 372570
Conc: 3.47 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.282 min
Delta R.T.: 0.006 min
Response: 50928
Conc: 0.25 ng/ml




#36 AR-1262-1

R.T.: 7.798 min
Delta R.T.: -0.004 min
Response: 194937
Conc: 2.12 ng/ml



#36 AR-1262-1


R.T.: 6.822 min
Delta R.T.: -0.003 min
Response: 34152
Conc: 0.55 ng/ml



#37 AR-1262-2


R.T.: 8.367 min
Delta R.T.: 0.013 min
Response: 372570
Conc: 2.83 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 7.036 min
Delta R.T.: 0.004 min
Response: 154792
Conc: 1.23 ng/ml




#38 AR-1262-3

R.T.: 8.678 min
Delta R.T.: 0.006 min
Response: 59269
Conc: 0.63 ng/ml



#38 AR-1262-3


R.T.: 7.563 min
Delta R.T.: 0.002 min
Response: 165043
Conc: 1.66 ng/ml



#39 AR-1262-4


R.T.: 8.785 min
Delta R.T.: 0.023 min
Response: 6815
Conc: 0.09 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.625 min
Delta R.T.: 0.000 min
Response: 162184
Conc: 0.90 ng/ml




#40 AR-1262-5

R.T.: 9.433 min
Delta R.T.: 0.001 min
Response: 9550
Conc: 0.22 ng/ml



#40 AR-1262-5


R.T.: 0.000 min
Exp R.T. : 8.124 min
Response: 0
Conc: N.D.



#41 AR-1268-1


R.T.: 8.665 min
Delta R.T.: -0.004 min
Response: 47995
Conc: 0.30 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.563 min
Delta R.T.: 0.002 min
Response: 165043
Conc: 0.60 ng/ml




#42 AR-1268-2

R.T.: 8.785 min
Delta R.T.: 0.019 min
Response: 6815
Conc: 0.05 ng/ml



#42 AR-1268-2


R.T.: 7.625 min
Delta R.T.: -0.002 min
Response: 162184
Conc: 0.64 ng/ml



#43 AR-1268-3


R.T.: 8.997 min
Delta R.T.: -0.001 min
Response: 194031
Conc: 1.50 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.833 min
Delta R.T.: 0.000 min
Response: 68349
Conc: 0.32 ng/ml




#44 AR-1268-4

R.T.: 9.433 min
Delta R.T.: 0.000 min
Response: 9550
Conc: 0.20 ng/ml



#44 AR-1268-4


R.T.: 0.000 min
Exp R.T. : 8.124 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 9.885 min
Delta R.T.: 0.027 min
Response: 90213
Conc: 0.27 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.425 min
Delta R.T.: 0.006 min
Response: 68054
Conc: 0.12 ng/ml