

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060922\
 Data File : PP048576.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Jun 2022 09:59
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 08 14:59:44 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jun 07 11:09:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.034	3.202	119.1E6	100.7E6	49.719	52.907
2) SA Decachlor...	10.384	8.619	90695946	59119386	47.584	48.278
Target Compounds						
26) L6 AR-1254-1	6.263	5.221	42775568	48431501	459.955	479.095
27) L6 AR-1254-2	6.506	5.382	62572445	41924930	451.100	476.138
28) L6 AR-1254-3	6.913	5.817	63520129	65217966	446.752	469.165
29) L6 AR-1254-4	7.232	6.067	46183567	40638496	433.387	468.636
30) L6 AR-1254-5	7.698	6.521	51071622	56419112	430.216	460.018



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060922\
Data File : PP048576.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Jun 2022 09:59
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 08 14:59:44 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060722.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jun 07 11:09:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.034 min
Delta R.T.: 0.000 min
Response: 119136635
Conc: 49.72 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.202 min
Delta R.T.: -0.002 min
Response: 100701500
Conc: 52.91 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.384 min
Delta R.T.: -0.003 min
Response: 90695946
Conc: 47.58 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.619 min
Delta R.T.: -0.009 min
Response: 59119386
Conc: 48.28 ng/ml



#26 AR-1254-1


R.T.: 6.263 min
Delta R.T.: -0.001 min
Response: 42775568
Conc: 459.96 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.221 min
Delta R.T.: -0.005 min
Response: 48431501
Conc: 479.09 ng/ml




#27 AR-1254-2

R.T.: 6.506 min
Delta R.T.: 0.000 min
Response: 62572445
Conc: 451.10 ng/ml



#27 AR-1254-2


R.T.: 5.382 min
Delta R.T.: -0.005 min
Response: 41924930
Conc: 476.14 ng/ml



#28 AR-1254-3


R.T.: 6.913 min
Delta R.T.: -0.002 min
Response: 63520129
Conc: 446.75 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 5.817 min
Delta R.T.: -0.005 min
Response: 65217966
Conc: 469.17 ng/ml




#29 AR-1254-4

R.T.: 7.232 min
Delta R.T.: -0.001 min
Response: 46183567
Conc: 433.39 ng/ml



#29 AR-1254-4


R.T.: 6.067 min
Delta R.T.: -0.006 min
Response: 40638496
Conc: 468.64 ng/ml



#30 AR-1254-5

R.T.: 7.698 min
Delta R.T.: 0.000 min
Response: 51071622
Conc: 430.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.521 min
Delta R.T.: -0.006 min
Response: 56419112
Conc: 460.02 ng/ml