

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061022\
 Data File : PP048651.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Jun 2022 17:56
 Operator : YP\AJ
 Sample : N3266-01
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 OH-WATER-0608

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/10/2022
 Supervised By :Ankita Jodhani 06/10/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 10 03:11:47 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jun 07 11:09:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.037	3.202	55250150	48062648	23.057	25.252m
2) SA Decachlor...	10.384	8.617	44244486	28939015	23.213	23.632

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061022\
Data File : PP048651.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jun 2022 17:56
Operator : YP\AJ
Sample : N3266-01
Misc :
ALS Vial : 25 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
OH-WATER-0608


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/10/2022
Supervised By :Ankita Jodhani 06/10/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 10 03:11:47 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060722.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jun 07 11:09:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.037 min
Delta R.T.: 0.002 min
Response: 55250150
Conc: 23.06 ng/ml

Instrument :
ECD_P
ClientSampleId :
OH-WATER-0608


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/10/2022
Supervised By :Ankita Jodhani 06/10/2022




#1 Tetrachloro-m-xylene

R.T.: 3.202 min
Delta R.T.: -0.002 min
Response: 48062648
Conc: 25.25 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.384 min
Delta R.T.: -0.002 min
Response: 44244486
Conc: 23.21 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.617 min
Delta R.T.: -0.011 min
Response: 28939015
Conc: 23.63 ng/ml