

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061022\
 Data File : PP048653.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Jun 2022 18:41
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 10 03:12:49 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jun 07 11:09:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.037	3.202	121.2E6	104.3E6	50.589	54.778
2) SA Decachlor...	10.386	8.618	95618521	62057664	50.167	50.677
Target Compounds						
16) L4 AR-1242-1	5.325	4.361	31312704	28461744	463.815	511.561
17) L4 AR-1242-2	5.348	4.380	45059995	38767193	463.668	510.625
18) L4 AR-1242-3	5.415	4.566	27803252	21927964	458.607	519.479
19) L4 AR-1242-4	5.525	4.658	22859943	20854923	449.403	518.235
20) L4 AR-1242-5	6.339	5.221	23817082	25193258	446.019	510.064



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061022\
Data File : PP048653.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jun 2022 18:41
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 10 03:12:49 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060722.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jun 07 11:09:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.037 min
Delta R.T.: 0.002 min
Response: 121220233
Conc: 50.59 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.202 min
Delta R.T.: -0.002 min
Response: 104260965
Conc: 54.78 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.386 min
Delta R.T.: 0.000 min
Response: 95618521
Conc: 50.17 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.618 min
Delta R.T.: -0.010 min
Response: 62057664
Conc: 50.68 ng/ml



#16 AR-1242-1


R.T.: 5.325 min
Delta R.T.: 0.001 min
Response: 31312704
Conc: 463.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.361 min
Delta R.T.: -0.004 min
Response: 28461744
Conc: 511.56 ng/ml




#17 AR-1242-2

R.T.: 5.348 min
Delta R.T.: 0.001 min
Response: 45059995
Conc: 463.67 ng/ml



#17 AR-1242-2


R.T.: 4.380 min
Delta R.T.: -0.005 min
Response: 38767193
Conc: 510.62 ng/ml



#18 AR-1242-3


R.T.: 5.415 min
Delta R.T.: 0.000 min
Response: 27803252
Conc: 458.61 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.566 min
Delta R.T.: -0.005 min
Response: 21927964
Conc: 519.48 ng/ml




#19 AR-1242-4

R.T.: 5.525 min
Delta R.T.: 0.001 min
Response: 22859943
Conc: 449.40 ng/ml



#19 AR-1242-4


R.T.: 4.658 min
Delta R.T.: -0.005 min
Response: 20854923
Conc: 518.24 ng/ml



#20 AR-1242-5

R.T.: 6.339 min
Delta R.T.: 0.000 min
Response: 23817082
Conc: 446.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.221 min
Delta R.T.: -0.006 min
Response: 25193258
Conc: 510.06 ng/ml