

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061022\
 Data File : PP048662.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Jun 2022 20:58
 Operator : YP\AJ
 Sample : N3245-04 10X
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 SOM-2022-36B

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/10/2022
 Supervised By :Ankita Jodhani 06/10/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 10 03:17:15 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jun 07 11:09:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.037	3.202	3929375	3973488	1.640	2.088 #
2) SA Decachlor...	10.389	8.622	3541464	1966803	1.858	1.606m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061022\
Data File : PP048662.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jun 2022 20:58
Operator : YP\AJ
Sample : N3245-04 10X
Misc :
ALS Vial : 31 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
SOM-2022-36B


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/10/2022
Supervised By :Ankita Jodhani 06/10/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 10 03:17:15 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060722.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jun 07 11:09:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





#1 Tetrachloro-m-xylene

R.T.: 4.037 min
Delta R.T.: 0.002 min
Response: 3929375
Conc: 1.64 ng/ml


Instrument :

ECD_P

ClientSampleId :
SOM-2022-36B


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/10/2022
Supervised By :Ankita Jodhani 06/10/2022




#1 Tetrachloro-m-xylene

R.T.: 3.202 min
Delta R.T.: -0.003 min
Response: 3973488
Conc: 2.09 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.389 min
Delta R.T.: 0.002 min
Response: 3541464
Conc: 1.86 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.622 min
Delta R.T.: -0.006 min
Response: 1966803
Conc: 1.61 ng/ml m