

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061022\
 Data File : PP048663.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Jun 2022 21:14
 Operator : YP\AJ
 Sample : PB145443BL
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PB145443BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 10 03:17:45 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jun 07 11:09:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.037	3.202	39635617	40315769	16.541	21.181 #
2) SA Decachlor...	10.386	8.617	39224236	25341073	20.579	20.694

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061022\
Data File : PP048663.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jun 2022 21:14
Operator : YP\AJ
Sample : PB145443BL
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB145443BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 10 03:17:45 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060722.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jun 07 11:09:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.037 min
Delta R.T.: 0.002 min
Response: 39635617
Conc: 16.54 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB145443BL




#1 Tetrachloro-m-xylene

R.T.: 3.202 min
Delta R.T.: -0.002 min
Response: 40315769
Conc: 21.18 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.386 min
Delta R.T.: 0.000 min
Response: 39224236
Conc: 20.58 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.617 min
Delta R.T.: -0.011 min
Response: 25341073
Conc: 20.69 ng/ml