

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061022\
 Data File : PP048664.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Jun 2022 21:29
 Operator : YP\AJ
 Sample : PB14543BS
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PB14543BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 10 03:18:18 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jun 07 11:09:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.053	3.202	36195804	34202257	15.106	17.970
2) SA Decachlor...	10.403	8.616	31607367	22163668	16.583	18.099
Target Compounds						
3) L1 AR-1016-1	5.341f	4.361	26223458	25409307	304.796	366.768
4) L1 AR-1016-2	5.341	4.380	26223458	34477267	215.388	363.500 #
5) L1 AR-1016-3	5.431f	4.566	23370656	19174551	308.342	368.658
6) L1 AR-1016-4	5.541f	4.615	19222645	14902574	305.583	370.232
7) L1 AR-1016-5	5.865f	4.841	19220142	19040119	311.331	373.347
31) L7 AR-1260-1	7.111f	5.958	34621617	32583160	332.274	386.143
32) L7 AR-1260-2	7.399f	6.164	41704065	39074162	342.492	381.475
33) L7 AR-1260-3	7.796f	6.329	27282401	35034029	322.259	361.504
34) L7 AR-1260-4	8.048f	6.845	35225705	24739873	340.469	349.544
35) L7 AR-1260-5	8.410f	7.110	65977440	55264033	337.037	344.127



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061022\
Data File : PP048664.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jun 2022 21:29
Operator : YP\AJ
Sample : PB14543BS
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB14543BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 10 03:18:18 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060722.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jun 07 11:09:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.053 min
Delta R.T.: 0.018 min
Response: 36195804
Conc: 15.11 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB14543BS




#1 Tetrachloro-m-xylene

R.T.: 3.202 min
Delta R.T.: -0.003 min
Response: 34202257
Conc: 17.97 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.403 min
Delta R.T.: 0.016 min
Response: 31607367
Conc: 16.58 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.616 min
Delta R.T.: -0.011 min
Response: 22163668
Conc: 18.10 ng/ml



#3 AR-1016-1


R.T.: 5.341 min
Delta R.T.: 0.018 min
Response: 26223458
Conc: 304.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB14543BS




#3 AR-1016-1

R.T.: 4.361 min
Delta R.T.: -0.004 min
Response: 25409307
Conc: 366.77 ng/ml




#4 AR-1016-2

R.T.: 5.341 min
Delta R.T.: -0.007 min
Response: 26223458
Conc: 215.39 ng/ml



#4 AR-1016-2


R.T.: 4.380 min
Delta R.T.: -0.005 min
Response: 34477267
Conc: 363.50 ng/ml



#5 AR-1016-3


R.T.: 5.431 min
Delta R.T.: 0.017 min
Response: 23370656
Conc: 308.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB14543BS




#5 AR-1016-3

R.T.: 4.566 min
Delta R.T.: -0.005 min
Response: 19174551
Conc: 368.66 ng/ml




#6 AR-1016-4

R.T.: 5.541 min
Delta R.T.: 0.018 min
Response: 19222645
Conc: 305.58 ng/ml



#6 AR-1016-4


R.T.: 4.615 min
Delta R.T.: -0.005 min
Response: 14902574
Conc: 370.23 ng/ml



#7 AR-1016-5


R.T.: 5.865 min
Delta R.T.: 0.017 min
Response: 19220142
Conc: 311.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB14543BS




#7 AR-1016-5

R.T.: 4.841 min
Delta R.T.: -0.006 min
Response: 19040119
Conc: 373.35 ng/ml




#31 AR-1260-1

R.T.: 7.111 min
Delta R.T.: 0.017 min
Response: 34621617
Conc: 332.27 ng/ml



#31 AR-1260-1


R.T.: 5.958 min
Delta R.T.: -0.008 min
Response: 32583160
Conc: 386.14 ng/ml



#32 AR-1260-2


R.T.: 7.399 min
Delta R.T.: 0.017 min
Response: 41704065
Conc: 342.49 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB14543BS




#32 AR-1260-2

R.T.: 6.164 min
Delta R.T.: -0.008 min
Response: 39074162
Conc: 381.47 ng/ml




#33 AR-1260-3

R.T.: 7.796 min
Delta R.T.: 0.017 min
Response: 27282401
Conc: 322.26 ng/ml



#33 AR-1260-3


R.T.: 6.329 min
Delta R.T.: -0.008 min
Response: 35034029
Conc: 361.50 ng/ml



#34 AR-1260-4


R.T.: 8.048 min
Delta R.T.: 0.017 min
Response: 35225705
Conc: 340.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB14543BS




#34 AR-1260-4

R.T.: 6.845 min
Delta R.T.: -0.008 min
Response: 24739873
Conc: 349.54 ng/ml



#35 AR-1260-5

R.T.: 8.410 min
Delta R.T.: 0.017 min
Response: 65977440
Conc: 337.04 ng/ml



#35 AR-1260-5

R.T.: 7.110 min
Delta R.T.: -0.008 min
Response: 55264033
Conc: 344.13 ng/ml