

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061022\
 Data File : PP048677.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Jun 2022 01:15
 Operator : YP\AJ
 Sample : N3245-10 10X
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 SOM-2022-39B

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/10/2022
 Supervised By :Ankita Jodhani 06/10/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 10 03:24:35 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jun 07 11:09:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.037	3.202	3385082	3766326	1.413	1.979 #
2) SA Decachlor...	10.389	8.619	3501609	2023817	1.837	1.653m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061022\
Data File : PP048677.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jun 2022 01:15
Operator : YP\AJ
Sample : N3245-10 10X
Misc :
ALS Vial : 41 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
SOM-2022-39B


Manual Integrations
APPROVED

Reviewed By : Yogesh Patel 06/10/2022
Supervised By : Ankita Jodhani 06/10/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 10 03:24:35 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060722.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jun 07 11:09:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





#1 Tetrachloro-m-xylene

R.T.: 4.037 min
Delta R.T.: 0.002 min
Response: 3385082
Conc: 1.41 ng/ml


Instrument :

ECD_P

ClientSampleId :
SOM-2022-39B


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/10/2022
Supervised By :Ankita Jodhani 06/10/2022




#1 Tetrachloro-m-xylene

R.T.: 3.202 min
Delta R.T.: -0.003 min
Response: 3766326
Conc: 1.98 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.389 min
Delta R.T.: 0.002 min
Response: 3501609
Conc: 1.84 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.619 min
Delta R.T.: -0.009 min
Response: 2023817
Conc: 1.65 ng/ml m