

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061024\
 Data File : PP065540.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Jun 2024 16:08
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 11 01:12:39 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060524.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 06 05:25:59 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.825	4.070	93611954	137.7E6	49.349	49.329
2) SA Decachlor...	10.843	9.259	87025568	64165750	44.246	43.556
Target Compounds						
21) L5 AR-1248-1	6.003	5.184	19210771	26939495	472.365	504.730
22) L5 AR-1248-2	6.277	5.424	30093880	35915403	482.868	487.950
23) L5 AR-1248-3	6.485	5.468	32221541	37353589	478.490	488.807
24) L5 AR-1248-4	6.889	5.645	33627793	42266243	472.332	480.765
25) L5 AR-1248-5	6.930	6.044	32364673	36646222	466.880	459.357



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061024\
Data File : PP065540.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jun 2024 16:08
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 11 01:12:39 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060524.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 06 05:25:59 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.825 min
Delta R.T.: 0.000 min
Response: 93611954
Conc: 49.35 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.070 min
Delta R.T.: -0.001 min
Response: 137743091
Conc: 49.33 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.843 min
Delta R.T.: 0.000 min
Response: 87025568
Conc: 44.25 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.259 min
Delta R.T.: -0.002 min
Response: 64165750
Conc: 43.56 ng/ml



#21 AR-1248-1


R.T.: 6.003 min
Delta R.T.: 0.000 min
Response: 19210771
Conc: 472.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 5.184 min
Delta R.T.: 0.000 min
Response: 26939495
Conc: 504.73 ng/ml




#22 AR-1248-2

R.T.: 6.277 min
Delta R.T.: 0.000 min
Response: 30093880
Conc: 482.87 ng/ml



#22 AR-1248-2


R.T.: 5.424 min
Delta R.T.: 0.000 min
Response: 35915403
Conc: 487.95 ng/ml



#23 AR-1248-3


R.T.: 6.485 min
Delta R.T.: 0.000 min
Response: 32221541
Conc: 478.49 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.468 min
Delta R.T.: -0.001 min
Response: 37353589
Conc: 488.81 ng/ml




#24 AR-1248-4

R.T.: 6.889 min
Delta R.T.: 0.000 min
Response: 33627793
Conc: 472.33 ng/ml



#24 AR-1248-4


R.T.: 5.645 min
Delta R.T.: 0.000 min
Response: 42266243
Conc: 480.76 ng/ml



#25 AR-1248-5

R.T.: 6.930 min
Delta R.T.: 0.000 min
Response: 32364673
Conc: 466.88 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 6.044 min
Delta R.T.: 0.000 min
Response: 36646222
Conc: 459.36 ng/ml