

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061025\
 Data File : PP072790.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Jun 2025 18:53
 Operator : YP\AJ
 Sample : PB168398BL
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PB168398BL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/11/2025
 Supervised By :mohammad ahmed 06/12/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 11 06:19:28 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 24 03:32:20 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.497	3.793	42421603	37163574	21.097	23.251
2) SA Decachlor...	10.195	8.807	36195146	26772054	22.218m	25.290m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061025\
Data File : PP072790.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jun 2025 18:53
Operator : YP\AJ
Sample : PB168398BL
Misc :
ALS Vial : 16 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
PB168398BL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/11/2025
Supervised By :mohammad ahmed 06/12/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 11 06:19:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.497 min
Delta R.T.: 0.000 min
Response: 42421603
Conc: 21.10 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB168398BL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/11/2025
Supervised By :mohammad ahmed 06/12/2025




#1 Tetrachloro-m-xylene

R.T.: 3.793 min
Delta R.T.: 0.000 min
Response: 37163574
Conc: 23.25 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.195 min
Delta R.T.: -0.017 min
Response: 36195146
Conc: 22.22 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.807 min
Delta R.T.: -0.015 min
Response: 26772054
Conc: 25.29 ng/ml m