

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061121\
 Data File : PP036352.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 11 Jun 2021 20:50
 Operator : DD\AJ
 Sample : AR1248ICC1000
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 12 01:05:45 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061121.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jun 12 01:05:23 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.965	3.959	3954169	2948702	95.891	98.162
2) SA Decachlor...	10.998	9.266	3514126	2985327	91.858	93.550
Target Compounds						
21) L5 AR-1248-1	6.285	5.215	827254	606867	903.866	934.907
22) L5 AR-1248-2	6.579	5.479	1049677	782825	899.329	920.363
23) L5 AR-1248-3	6.797	5.523	1276928	803177	904.330	920.129
24) L5 AR-1248-4	7.225	5.706	1424005	999603	908.044	931.410
25) L5 AR-1248-5	7.264	6.130	1392125	1018596	905.077	934.792



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061121\
Data File : PP036352.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 11 Jun 2021 20:50
Operator : DD\AJ
Sample : AR1248ICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 12 01:05:45 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061121.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jun 12 01:05:23 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.965 min
Delta R.T.: 0.000 min
Response: 3954169
Conc: 95.89 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.959 min
Delta R.T.: 0.000 min
Response: 2948702
Conc: 98.16 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.998 min
Delta R.T.: -0.001 min
Response: 3514126
Conc: 91.86 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.266 min
Delta R.T.: 0.000 min
Response: 2985327
Conc: 93.55 ng/ml



#21 AR-1248-1


R.T.: 6.285 min
Delta R.T.: 0.000 min
Response: 827254
Conc: 903.87 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 5.215 min
Delta R.T.: 0.000 min
Response: 606867
Conc: 934.91 ng/ml




#22 AR-1248-2

R.T.: 6.579 min
Delta R.T.: 0.000 min
Response: 1049677
Conc: 899.33 ng/ml



#22 AR-1248-2


R.T.: 5.479 min
Delta R.T.: 0.000 min
Response: 782825
Conc: 920.36 ng/ml



#23 AR-1248-3


R.T.: 6.797 min
Delta R.T.: 0.000 min
Response: 1276928
Conc: 904.33 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.523 min
Delta R.T.: 0.000 min
Response: 803177
Conc: 920.13 ng/ml




#24 AR-1248-4

R.T.: 7.225 min
Delta R.T.: 0.000 min
Response: 1424005
Conc: 908.04 ng/ml



#24 AR-1248-4


R.T.: 5.706 min
Delta R.T.: 0.000 min
Response: 999603
Conc: 931.41 ng/ml



#25 AR-1248-5

R.T.: 7.264 min
Delta R.T.: 0.000 min
Response: 1392125
Conc: 905.08 ng/ml

Instrument :
ECD_P
ClientSampleId :



#25 AR-1248-5

R.T.: 6.130 min
Delta R.T.: 0.000 min
Response: 1018596
Conc: 934.79 ng/ml