

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061121\
 Data File : PP036360.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 11 Jun 2021 23:04
 Operator : DD\AJ
 Sample : AR1254ICC250
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 12 01:26:41 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061121.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jun 12 01:25:10 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.964	3.959	1065239	754262	25.648	25.183
2)	SA Decachlor...	10.998	9.266	1039944	841202	26.504	25.658
Target Compounds							
26)	L6 AR-1254-1	7.194	6.087	435353	425021	274.390	264.744
27)	L6 AR-1254-2	7.424	6.247	663298	384956	271.954	265.621
28)	L6 AR-1254-3	7.805	6.668	709487	615521	270.169	258.019
29)	L6 AR-1254-4	8.101	6.907	538816	407746	269.565	263.669
30)	L6 AR-1254-5	8.530	7.338	597574	571773	266.570	257.288



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061121\
Data File : PP036360.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 11 Jun 2021 23:04
Operator : DD\AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 12 01:26:41 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061121.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jun 12 01:25:10 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.964 min
Delta R.T.: -0.002 min
Response: 1065239
Conc: 25.65 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.959 min
Delta R.T.: -0.002 min
Response: 754262
Conc: 25.18 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.998 min
Delta R.T.: -0.001 min
Response: 1039944
Conc: 26.50 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.266 min
Delta R.T.: -0.001 min
Response: 841202
Conc: 25.66 ng/ml



#26 AR-1254-1


R.T.: 7.194 min
Delta R.T.: -0.001 min
Response: 435353
Conc: 274.39 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250




#26 AR-1254-1

R.T.: 6.087 min
Delta R.T.: -0.001 min
Response: 425021
Conc: 264.74 ng/ml




#27 AR-1254-2

R.T.: 7.424 min
Delta R.T.: -0.001 min
Response: 663298
Conc: 271.95 ng/ml



#27 AR-1254-2


R.T.: 6.247 min
Delta R.T.: -0.002 min
Response: 384956
Conc: 265.62 ng/ml



#28 AR-1254-3


R.T.: 7.805 min
Delta R.T.: -0.002 min
Response: 709487
Conc: 270.17 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250




#28 AR-1254-3

R.T.: 6.668 min
Delta R.T.: -0.001 min
Response: 615521
Conc: 258.02 ng/ml




#29 AR-1254-4

R.T.: 8.101 min
Delta R.T.: -0.002 min
Response: 538816
Conc: 269.57 ng/ml



#29 AR-1254-4


R.T.: 6.907 min
Delta R.T.: -0.002 min
Response: 407746
Conc: 263.67 ng/ml



#30 AR-1254-5

R.T.: 8.530 min
Delta R.T.: -0.002 min
Response: 597574
Conc: 266.57 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250



#30 AR-1254-5

R.T.: 7.338 min
Delta R.T.: -0.002 min
Response: 571773
Conc: 257.29 ng/ml