

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061124\
 Data File : PP065595.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jun 2024 16:55
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 12 00:14:37 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060524.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 06 05:25:59 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.824	4.068	92336854	135.8E6	48.677	48.637
2) SA Decachlor...	10.837	9.255	90223961	65517866	45.872	44.473
Target Compounds						
16) L4 AR-1242-1	6.003	5.182	25664845	33556655	469.338	466.970
17) L4 AR-1242-2	6.025	5.203	37699789	46810658	475.399	470.711
18) L4 AR-1242-3	6.089	5.384	24641706	24802546	487.489	461.109
19) L4 AR-1242-4	6.189	5.466	19753094	24774132	481.367	465.153
20) L4 AR-1242-5	6.928	5.998	19608488	27044725	449.510	455.158



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061124\
Data File : PP065595.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2024 16:55
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 12 00:14:37 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060524.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 06 05:25:59 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.824 min
Delta R.T.: -0.002 min
Response: 92336854
Conc: 48.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.068 min
Delta R.T.: -0.003 min
Response: 135810525
Conc: 48.64 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.837 min
Delta R.T.: -0.006 min
Response: 90223961
Conc: 45.87 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.255 min
Delta R.T.: -0.006 min
Response: 65517866
Conc: 44.47 ng/ml



#16 AR-1242-1


R.T.: 6.003 min
Delta R.T.: -0.002 min
Response: 25664845
Conc: 469.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 5.182 min
Delta R.T.: -0.004 min
Response: 33556655
Conc: 466.97 ng/ml




#17 AR-1242-2

R.T.: 6.025 min
Delta R.T.: -0.002 min
Response: 37699789
Conc: 475.40 ng/ml



#17 AR-1242-2


R.T.: 5.203 min
Delta R.T.: -0.004 min
Response: 46810658
Conc: 470.71 ng/ml



#18 AR-1242-3


R.T.: 6.089 min
Delta R.T.: -0.003 min
Response: 24641706
Conc: 487.49 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 5.384 min
Delta R.T.: -0.004 min
Response: 24802546
Conc: 461.11 ng/ml




#19 AR-1242-4

R.T.: 6.189 min
Delta R.T.: -0.002 min
Response: 19753094
Conc: 481.37 ng/ml



#19 AR-1242-4


R.T.: 5.466 min
Delta R.T.: -0.004 min
Response: 24774132
Conc: 465.15 ng/ml



#20 AR-1242-5

R.T.: 6.928 min
Delta R.T.: -0.002 min
Response: 19608488
Conc: 449.51 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.998 min
Delta R.T.: -0.004 min
Response: 27044725
Conc: 455.16 ng/ml