

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061225\
 Data File : PP072875.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Jun 2025 16:33
 Operator : YP\AJ
 Sample : Q2295-05
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 TP02-MH5-WC

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/13/2025
 Supervised By :mohammad ahmed 06/16/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 12 22:47:17 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 24 03:32:20 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.496	3.793	41419878	36045841	20.599	22.552
2) SA Decachlor...	10.193	8.804	29721813	22824432	18.244m	21.561m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061225\
Data File : PP072875.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Jun 2025 16:33
Operator : YP\AJ
Sample : Q2295-05
Misc :
ALS Vial : 13 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
TP02-MH5-WC


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/13/2025
Supervised By :mohammad ahmed 06/16/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 12 22:47:17 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.496 min
Delta R.T.: -0.002 min
Response: 41419878
Conc: 20.60 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP02-MH5-WC


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/13/2025
Supervised By :mohammad ahmed 06/16/2025




#1 Tetrachloro-m-xylene

R.T.: 3.793 min
Delta R.T.: 0.000 min
Response: 36045841
Conc: 22.55 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.193 min
Delta R.T.: -0.019 min
Response: 29721813
Conc: 18.24 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.804 min
Delta R.T.: -0.017 min
Response: 22824432
Conc: 21.56 ng/ml m