

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061321\
 Data File : PP036399.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 13 Jun 2021 13:56
 Operator : DD\AJ
 Sample : M2676-01
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 14 04:47:51 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061121.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jun 12 04:04:46 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.965	3.959	770234	546934	18.242	18.209
2)	SA Decachlor...	10.998	9.266	752261	575794	20.833	18.023
Target Compounds							
8)	L2 AR-1221-1	5.217	0.000	36320	0	73.029	N.D. #
9)	L2 AR-1221-2	5.315	0.000	12691	0	32.904	N.D. #
12)	L3 AR-1232-2	6.004	0.000	21973	0	45.964	N.D. #
28)	L6 AR-1254-3	7.818	0.000	65072	0	23.445	N.D. #
30)	L6 AR-1254-5	8.533	0.000	4087	0	1.755	N.D. #
33)	L7 AR-1260-3	8.600	0.000	3724	0	2.130	N.D. #
36)	L8 AR-1262-1	8.600	0.000	3724	0	1.385	N.D. #
38)	L8 AR-1262-3	0.000	8.183	0	37508	N.D.	21.893 #
39)	L8 AR-1262-4	0.000	8.251	0	43554	N.D.	13.883 #
41)	L9 AR-1268-1	0.000	8.183	0	37508	N.D.	8.123 #
42)	L9 AR-1268-2	0.000	8.251	0	43554	N.D.	10.125 #
44)	L9 AR-1268-4	10.279f	0.000	13447	0	6.869	N.D. #
45)	L9 AR-1268-5	0.000	9.041	0	5327	N.D.	0.433 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061321\
Data File : PP036399.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 13 Jun 2021 13:56
Operator : DD\AJ
Sample : M2676-01
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 14 04:47:51 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061121.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jun 12 04:04:46 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.965 min
Delta R.T.: 0.000 min
Response: 770234
Conc: 18.24 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.959 min
Delta R.T.: 0.000 min
Response: 546934
Conc: 18.21 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.998 min
Delta R.T.: 0.001 min
Response: 752261
Conc: 20.83 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.266 min
Delta R.T.: -0.001 min
Response: 575794
Conc: 18.02 ng/ml



#8 AR-1221-1


R.T.: 5.217 min
Delta R.T.: 0.001 min
Response: 36320
Conc: 73.03 ng/ml

Instrument :
ECD_P
ClientSampleId :




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T. : 4.218 min
Response: 0
Conc: N.D.




#9 AR-1221-2

R.T.: 5.315 min
Delta R.T.: 0.001 min
Response: 12691
Conc: 32.90 ng/ml



#9 AR-1221-2


R.T.: 0.000 min
Exp R.T. : 4.316 min
Response: 0
Conc: N.D.



#12 AR-1232-2


R.T.: 6.004 min
Delta R.T.: 0.015 min
Response: 21973
Conc: 45.96 ng/ml

Instrument :
ECD_P
ClientSampleId :




#12 AR-1232-2

R.T.: 0.000 min
Exp R.T. : 5.235 min
Response: 0
Conc: N.D.




#28 AR-1254-3

R.T.: 7.818 min
Delta R.T.: 0.013 min
Response: 65072
Conc: 23.44 ng/ml



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T. : 6.667 min
Response: 0
Conc: N.D.



#30 AR-1254-5


R.T.: 8.533 min
Delta R.T.: 0.003 min
Response: 4087
Conc: 1.76 ng/ml

Instrument :
ECD_P
ClientSampleId :




#30 AR-1254-5

R.T.: 0.000 min
Exp R.T. : 7.338 min
Response: 0
Conc: N.D.




#33 AR-1260-3

R.T.: 8.600 min
Delta R.T.: -0.002 min
Response: 3724
Conc: 2.13 ng/ml



#33 AR-1260-3


R.T.: 0.000 min
Exp R.T. : 7.158 min
Response: 0
Conc: N.D.



#36 AR-1262-1


R.T.: 8.600 min
Delta R.T.: -0.004 min
Response: 3724
Conc: 1.39 ng/ml

Instrument :
ECD_P
ClientSampleId :




#36 AR-1262-1

R.T.: 0.000 min
Exp R.T. : 7.338 min
Response: 0
Conc: N.D.




#38 AR-1262-3

R.T.: 0.000 min
Exp R.T. : 9.477 min
Response: 0
Conc: N.D.



#38 AR-1262-3


R.T.: 8.183 min
Delta R.T.: 0.002 min
Response: 37508
Conc: 21.89 ng/ml



#39 AR-1262-4


R.T.: 0.000 min
Exp R.T. : 9.571 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 8.251 min
Delta R.T.: 0.008 min
Response: 43554
Conc: 13.88 ng/ml




#41 AR-1268-1

R.T.: 0.000 min
Exp R.T. : 9.488 min
Response: 0
Conc: N.D.



#41 AR-1268-1


R.T.: 8.183 min
Delta R.T.: -0.002 min
Response: 37508
Conc: 8.12 ng/ml



#42 AR-1268-2


R.T.: 0.000 min
Exp R.T.: 9.586 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#42 AR-1268-2

R.T.: 8.251 min
Delta R.T.: 0.002 min
Response: 43554
Conc: 10.13 ng/ml




#44 AR-1268-4

R.T.: 10.279 min
Delta R.T.: 0.022 min
Response: 13447
Conc: 6.87 ng/ml



#44 AR-1268-4


R.T.: 0.000 min
Exp R.T.: 8.747 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T.: 10.676 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 9.041 min
Delta R.T.: 0.012 min
Response: 5327
Conc: 0.43 ng/ml