

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061321\
 Data File : PP036417.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 13 Jun 2021 20:05
 Operator : DD\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 14 04:54:55 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061121.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jun 12 04:04:46 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.965	3.959	2050860	1476147	48.571	49.144
2) SA Decachlor...	10.999	9.265	1984774	1564514	54.965	48.972
Target Compounds						
26) L6 AR-1254-1	7.194	6.086	829313	806858	503.750	495.376
27) L6 AR-1254-2	7.424	6.247	1214048	720419	477.966	489.669
28) L6 AR-1254-3	7.805	6.666	1308529	1251001	471.450	520.390
29) L6 AR-1254-4	8.102	6.906	999307	770571	484.468	489.510
30) L6 AR-1254-5	8.530	7.336	1136151	1100940	487.896	497.950



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061321\
Data File : PP036417.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 13 Jun 2021 20:05
Operator : DD\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 14 04:54:55 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061121.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jun 12 04:04:46 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.965 min
Delta R.T.: 0.000 min
Response: 2050860
Conc: 48.57 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.959 min
Delta R.T.: 0.000 min
Response: 1476147
Conc: 49.14 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.999 min
Delta R.T.: 0.002 min
Response: 1984774
Conc: 54.96 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.265 min
Delta R.T.: -0.002 min
Response: 1564514
Conc: 48.97 ng/ml



#26 AR-1254-1


R.T.: 7.194 min
Delta R.T.: 0.000 min
Response: 829313
Conc: 503.75 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 6.086 min
Delta R.T.: 0.000 min
Response: 806858
Conc: 495.38 ng/ml




#27 AR-1254-2

R.T.: 7.424 min
Delta R.T.: 0.000 min
Response: 1214048
Conc: 477.97 ng/ml



#27 AR-1254-2


R.T.: 6.247 min
Delta R.T.: 0.000 min
Response: 720419
Conc: 489.67 ng/ml



#28 AR-1254-3


R.T.: 7.805 min
Delta R.T.: 0.000 min
Response: 1308529
Conc: 471.45 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.666 min
Delta R.T.: 0.000 min
Response: 1251001
Conc: 520.39 ng/ml




#29 AR-1254-4

R.T.: 8.102 min
Delta R.T.: 0.001 min
Response: 999307
Conc: 484.47 ng/ml



#29 AR-1254-4


R.T.: 6.906 min
Delta R.T.: -0.001 min
Response: 770571
Conc: 489.51 ng/ml



#30 AR-1254-5

R.T.: 8.530 min
Delta R.T.: 0.000 min
Response: 1136151
Conc: 487.90 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 7.336 min
Delta R.T.: -0.002 min
Response: 1100940
Conc: 497.95 ng/ml