

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061324\
 Data File : PP065648.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Jun 2024 10:11
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 13 23:33:08 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060524.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 06 05:25:59 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.823	4.067	98332770	141.7E6	51.838	50.731
2) SA Decachlor...	10.838	9.251	97069183	72950994	49.352	49.519
Target Compounds						
21) L5 AR-1248-1	6.001	5.180	20748145	27394359	510.167	513.252
22) L5 AR-1248-2	6.276	5.420	30986400	36614893	497.189	497.453
23) L5 AR-1248-3	6.483	5.464	33434722	38199937	496.506	499.883
24) L5 AR-1248-4	6.888	5.641	36786744	44026206	516.703	500.784
25) L5 AR-1248-5	6.928	6.040	36101027	40042343	520.779	501.927



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061324\
Data File : PP065648.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jun 2024 10:11
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 13 23:33:08 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060524.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 06 05:25:59 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.823 min
Delta R.T.: -0.003 min
Response: 98332770
Conc: 51.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.067 min
Delta R.T.: -0.004 min
Response: 141657273
Conc: 50.73 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.838 min
Delta R.T.: -0.005 min
Response: 97069183
Conc: 49.35 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.251 min
Delta R.T.: -0.009 min
Response: 72950994
Conc: 49.52 ng/ml



#21 AR-1248-1


R.T.: 6.001 min
Delta R.T.: -0.002 min
Response: 20748145
Conc: 510.17 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 5.180 min
Delta R.T.: -0.005 min
Response: 27394359
Conc: 513.25 ng/ml




#22 AR-1248-2

R.T.: 6.276 min
Delta R.T.: -0.002 min
Response: 30986400
Conc: 497.19 ng/ml



#22 AR-1248-2


R.T.: 5.420 min
Delta R.T.: -0.005 min
Response: 36614893
Conc: 497.45 ng/ml



#23 AR-1248-3


R.T.: 6.483 min
Delta R.T.: -0.002 min
Response: 33434722
Conc: 496.51 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.464 min
Delta R.T.: -0.005 min
Response: 38199937
Conc: 499.88 ng/ml




#24 AR-1248-4

R.T.: 6.888 min
Delta R.T.: -0.002 min
Response: 36786744
Conc: 516.70 ng/ml



#24 AR-1248-4


R.T.: 5.641 min
Delta R.T.: -0.005 min
Response: 44026206
Conc: 500.78 ng/ml



#25 AR-1248-5

R.T.: 6.928 min
Delta R.T.: -0.002 min
Response: 36101027
Conc: 520.78 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 6.040 min
Delta R.T.: -0.005 min
Response: 40042343
Conc: 501.93 ng/ml