

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061422\
 Data File : PP048789.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Jun 2022 03:29
 Operator : YP\AJ
 Sample : N3316-07 10X
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 FCA566M-A

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/15/2022
 Supervised By :Ankita Jodhani 06/15/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 15 05:20:36 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jun 07 11:09:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.037	3.201	3528515	3838465	1.473m	2.017m#
2) SA Decachlor...	10.390	8.611	3567798	2192641	1.872	1.791

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061422\
Data File : PP048789.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Jun 2022 03:29
Operator : YP\AJ
Sample : N3316-07 10X
Misc :
ALS Vial : 35 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
FCA566M-A


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/15/2022
Supervised By :Ankita Jodhani 06/15/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 15 05:20:36 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060722.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jun 07 11:09:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.037 min
Delta R.T.: 0.002 min
Response: 3528515
Conc: 1.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
FCA566M-A


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/15/2022
Supervised By :Ankita Jodhani 06/15/2022




#1 Tetrachloro-m-xylene

R.T.: 3.201 min
Delta R.T.: -0.004 min
Response: 3838465
Conc: 2.02 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.390 min
Delta R.T.: 0.003 min
Response: 3567798
Conc: 1.87 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.611 min
Delta R.T.: -0.016 min
Response: 2192641
Conc: 1.79 ng/ml