

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061424\
 Data File : PP065707.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Jun 2024 16:57
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 14 23:12:51 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060524.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 06 05:25:59 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.825	4.069	100.6E6	143.8E6	53.021	51.503
2) SA Decachlor...	10.841	9.253	100.2E6	73410098	50.935	49.831
Target Compounds						
16) L4 AR-1242-1	6.004	5.182	28419644	35543151	519.716	494.614
17) L4 AR-1242-2	6.027	5.202	41532626	49708531	523.732	499.851
18) L4 AR-1242-3	6.090	5.384	26422191	25113472	522.712	466.890
19) L4 AR-1242-4	6.190	5.466	21491188	26535153	523.723	498.217
20) L4 AR-1242-5	6.930	5.997	23517010	32053313	539.111	539.452



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061424\
Data File : PP065707.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Jun 2024 16:57
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 14 23:12:51 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060524.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 06 05:25:59 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.825 min
Delta R.T.: 0.000 min
Response: 100577956
Conc: 53.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.069 min
Delta R.T.: -0.002 min
Response: 143812486
Conc: 51.50 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.841 min
Delta R.T.: -0.001 min
Response: 100183009
Conc: 50.94 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.253 min
Delta R.T.: -0.007 min
Response: 73410098
Conc: 49.83 ng/ml



#16 AR-1242-1


R.T.: 6.004 min
Delta R.T.: 0.000 min
Response: 28419644
Conc: 519.72 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 5.182 min
Delta R.T.: -0.004 min
Response: 35543151
Conc: 494.61 ng/ml




#17 AR-1242-2

R.T.: 6.027 min
Delta R.T.: 0.000 min
Response: 41532626
Conc: 523.73 ng/ml



#17 AR-1242-2


R.T.: 5.202 min
Delta R.T.: -0.004 min
Response: 49708531
Conc: 499.85 ng/ml



#18 AR-1242-3


R.T.: 6.090 min
Delta R.T.: 0.000 min
Response: 26422191
Conc: 522.71 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 5.384 min
Delta R.T.: -0.005 min
Response: 25113472
Conc: 466.89 ng/ml




#19 AR-1242-4

R.T.: 6.190 min
Delta R.T.: 0.000 min
Response: 21491188
Conc: 523.72 ng/ml



#19 AR-1242-4


R.T.: 5.466 min
Delta R.T.: -0.004 min
Response: 26535153
Conc: 498.22 ng/ml



#20 AR-1242-5

R.T.: 6.930 min
Delta R.T.: 0.000 min
Response: 23517010
Conc: 539.11 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.997 min
Delta R.T.: -0.005 min
Response: 32053313
Conc: 539.45 ng/ml