

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061523\
 Data File : PP058486.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Jun 2023 23:49
 Operator : YP\AJ
 Sample : AR1242ICC050
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 16 04:24:49 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061523.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 16 03:42:54 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.457	3.769	10384901	7263295	4.350	4.286
2) SA Decachlor...	10.357	8.872	5853951	7252344	4.106	5.002
Target Compounds						
16) L4 AR-1242-1	5.640	4.870	2415406	2133562	47.589	48.682
17) L4 AR-1242-2	5.663	4.890	3475367	2718171	47.087	46.193
18) L4 AR-1242-3	5.726	5.068	2107837	1497430	46.235	45.067
19) L4 AR-1242-4	5.825	5.153	1764802	1657668	47.085	46.896
20) L4 AR-1242-5	6.573	5.681	1652358	1965075	47.232	46.346



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061523\
Data File : PP058486.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Jun 2023 23:49
Operator : YP\AJ
Sample : AR1242ICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 16 04:24:49 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061523.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 16 03:42:54 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.457 min
Delta R.T.: 0.000 min
Response: 10384901
Conc: 4.35 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050




#1 Tetrachloro-m-xylene

R.T.: 3.769 min
Delta R.T.: 0.000 min
Response: 7263295
Conc: 4.29 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.357 min
Delta R.T.: 0.000 min
Response: 5853951
Conc: 4.11 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.872 min
Delta R.T.: 0.000 min
Response: 7252344
Conc: 5.00 ng/ml



#16 AR-1242-1


R.T.: 5.640 min
Delta R.T.: 0.000 min
Response: 2415406
Conc: 47.59 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050




#16 AR-1242-1

R.T.: 4.870 min
Delta R.T.: 0.000 min
Response: 2133562
Conc: 48.68 ng/ml




#17 AR-1242-2

R.T.: 5.663 min
Delta R.T.: 0.000 min
Response: 3475367
Conc: 47.09 ng/ml



#17 AR-1242-2


R.T.: 4.890 min
Delta R.T.: 0.000 min
Response: 2718171
Conc: 46.19 ng/ml



#18 AR-1242-3


R.T.: 5.726 min
Delta R.T.: 0.000 min
Response: 2107837
Conc: 46.24 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050




#18 AR-1242-3

R.T.: 5.068 min
Delta R.T.: 0.000 min
Response: 1497430
Conc: 45.07 ng/ml




#19 AR-1242-4

R.T.: 5.825 min
Delta R.T.: 0.000 min
Response: 1764802
Conc: 47.09 ng/ml



#19 AR-1242-4


R.T.: 5.153 min
Delta R.T.: 0.000 min
Response: 1657668
Conc: 46.90 ng/ml



#20 AR-1242-5

R.T.: 6.573 min
Delta R.T.: 0.001 min
Response: 1652358
Conc: 47.23 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050



#20 AR-1242-5

R.T.: 5.681 min
Delta R.T.: 0.000 min
Response: 1965075
Conc: 46.35 ng/ml