

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061523\
 Data File : PP058477.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Jun 2023 21:20
 Operator : YP\AJ
 Sample : AR1660ICC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 16 03:21:02 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061523.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 16 03:20:37 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.458	3.770	112.4E6	81023249	50.000	50.000
2) SA Decachlor...	10.358	8.873	69882841	73100601	50.000	50.000
Target Compounds						
3) L1 AR-1016-1	5.641	4.871	31497265	27724683	500.000	500.000
4) L1 AR-1016-2	5.664	4.890	46409112	37525899	500.000	500.000
5) L1 AR-1016-3	5.727	5.069	28107771	20901574	500.000	500.000
6) L1 AR-1016-4	5.827	5.153	23127123	21743939	500.000	500.000
7) L1 AR-1016-5	6.126	5.327	23314615	22380168	500.000	500.000
31) L7 AR-1260-1	7.266	6.372	40871260	42591007	500.000	500.000
32) L7 AR-1260-2	7.526	6.559	45200264	50049991	500.000	500.000
33) L7 AR-1260-3	7.889	6.716	29619112	47301947	500.000	500.000
34) L7 AR-1260-4	8.119	7.192	35499765	34705409	500.000	500.000
35) L7 AR-1260-5	8.453	7.433	66770744	79187210	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061523\
Data File : PP058477.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Jun 2023 21:20
Operator : YP\AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 16 03:21:02 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061523.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 16 03:20:37 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.458 min
Delta R.T.: 0.000 min
Response: 112391783
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.770 min
Delta R.T.: 0.000 min
Response: 81023249
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.358 min
Delta R.T.: 0.000 min
Response: 69882841
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.873 min
Delta R.T.: 0.000 min
Response: 73100601
Conc: 50.00 ng/ml



#3 AR-1016-1


R.T.: 5.641 min
Delta R.T.: 0.000 min
Response: 31497265
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500




#3 AR-1016-1

R.T.: 4.871 min
Delta R.T.: 0.000 min
Response: 27724683
Conc: 500.00 ng/ml




#4 AR-1016-2

R.T.: 5.664 min
Delta R.T.: 0.000 min
Response: 46409112
Conc: 500.00 ng/ml



#4 AR-1016-2


R.T.: 4.890 min
Delta R.T.: 0.000 min
Response: 37525899
Conc: 500.00 ng/ml



#5 AR-1016-3


R.T.: 5.727 min
Delta R.T.: 0.000 min
Response: 28107771
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500




#5 AR-1016-3

R.T.: 5.069 min
Delta R.T.: 0.000 min
Response: 20901574
Conc: 500.00 ng/ml




#6 AR-1016-4

R.T.: 5.827 min
Delta R.T.: 0.000 min
Response: 23127123
Conc: 500.00 ng/ml



#6 AR-1016-4


R.T.: 5.153 min
Delta R.T.: 0.000 min
Response: 21743939
Conc: 500.00 ng/ml



#7 AR-1016-5


R.T.: 6.126 min
Delta R.T.: 0.000 min
Response: 23314615
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500




#7 AR-1016-5

R.T.: 5.327 min
Delta R.T.: 0.000 min
Response: 22380168
Conc: 500.00 ng/ml




#31 AR-1260-1

R.T.: 7.266 min
Delta R.T.: 0.000 min
Response: 40871260
Conc: 500.00 ng/ml



#31 AR-1260-1


R.T.: 6.372 min
Delta R.T.: 0.000 min
Response: 42591007
Conc: 500.00 ng/ml



#32 AR-1260-2


R.T.: 7.526 min
Delta R.T.: 0.000 min
Response: 45200264
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500




#32 AR-1260-2

R.T.: 6.559 min
Delta R.T.: 0.000 min
Response: 50049991
Conc: 500.00 ng/ml




#33 AR-1260-3

R.T.: 7.889 min
Delta R.T.: 0.000 min
Response: 29619112
Conc: 500.00 ng/ml



#33 AR-1260-3


R.T.: 6.716 min
Delta R.T.: 0.000 min
Response: 47301947
Conc: 500.00 ng/ml



#34 AR-1260-4


R.T.: 8.119 min
Delta R.T.: 0.000 min
Response: 35499765
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500




#34 AR-1260-4

R.T.: 7.192 min
Delta R.T.: 0.000 min
Response: 34705409
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 8.453 min
Delta R.T.: 0.000 min
Response: 66770744
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 7.433 min
Delta R.T.: 0.000 min
Response: 79187210
Conc: 500.00 ng/ml